

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060123\
 Data File : P0095421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 04:17
 Operator : YP/AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 06:27:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:15:34 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.436	3.655	16832644	7541856	4.647	4.610
2) SA Decachlor...	10.280	8.714	23209343	12439313	5.030	5.374
Target Compounds						
41) L9 AR-1268-1	8.728	7.594	18551404	11604811	50.302	53.691
42) L9 AR-1268-2	8.825	7.659	17488597	9904278	51.620	51.890
43) L9 AR-1268-3	9.061	7.868	14566045	8538363	50.383	52.673
44) L9 AR-1268-4	9.497	8.157	6165484	3594255	48.481	52.243
45) L9 AR-1268-5	9.926	8.453	46191023	23841966	49.107	49.706



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060123\
Data File : PO095421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 04:17
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 06:27:37 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:15:34 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.436 min
Delta R.T.: 0.000 min
Response: 16832644
Conc: 4.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.655 min
Delta R.T.: 0.000 min
Response: 7541856
Conc: 4.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.280 min
Delta R.T.: 0.000 min
Response: 23209343
Conc: 5.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.714 min
Delta R.T.: 0.000 min
Response: 12439313
Conc: 5.37 ng/ml



#41 AR-1268-1


R.T.: 8.728 min
Delta R.T.: 0.000 min
Response: 18551404
Conc: 50.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#41 AR-1268-1

R.T.: 7.594 min
Delta R.T.: 0.000 min
Response: 11604811
Conc: 53.69 ng/ml




#42 AR-1268-2

R.T.: 8.825 min
Delta R.T.: 0.000 min
Response: 17488597
Conc: 51.62 ng/ml



#42 AR-1268-2


R.T.: 7.659 min
Delta R.T.: 0.000 min
Response: 9904278
Conc: 51.89 ng/ml



#43 AR-1268-3


R.T.: 9.061 min
Delta R.T.: 0.000 min
Response: 14566045
Conc: 50.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#43 AR-1268-3

R.T.: 7.868 min
Delta R.T.: 0.000 min
Response: 8538363
Conc: 52.67 ng/ml




#44 AR-1268-4

R.T.: 9.497 min
Delta R.T.: 0.000 min
Response: 6165484
Conc: 48.48 ng/ml



#44 AR-1268-4


R.T.: 8.157 min
Delta R.T.: 0.000 min
Response: 3594255
Conc: 52.24 ng/ml



#45 AR-1268-5

R.T.: 9.926 min
Delta R.T.: 0.000 min
Response: 46191023
Conc: 49.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 8.453 min
Delta R.T.: 0.000 min
Response: 23841966
Conc: 49.71 ng/ml