

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0060123\
 Data File : P0095422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 04:34
 Operator : YP/AJ
 Sample : P0060123ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 ICVPO060123

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/02/2023
 Supervised By :Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 06:54:06 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:53:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.437	3.655	194.6E6	83214511	52.813	53.054
2)	SA Decachlor...	10.279	8.714	143.9E6	70838325	53.159	54.757
Target Compounds							
3)	L1 AR-1016-1	5.617	4.750	62671108	29893876	517.102	513.399
4)	L1 AR-1016-2	5.641	4.769	87367590	39842100	519.978	519.622
5)	L1 AR-1016-3	5.703	4.946	54451431	21619640	517.728	520.784
6)	L1 AR-1016-4	5.802	4.989	43802104	17595650	525.756	499.047
7)	L1 AR-1016-5	6.100	5.203	43903216	23502644	530.755	528.281m
31)	L7 AR-1260-1	7.236	6.246	76854304	45149302	518.327	529.238
32)	L7 AR-1260-2	7.495	6.435	90907387	52933231	517.906	518.032
33)	L7 AR-1260-3	7.857	6.590	58666352	48961800	522.605	522.928
34)	L7 AR-1260-4	8.084	7.065	71142818	35791122	522.123	526.174
35)	L7 AR-1260-5	8.411	7.308	133.3E6	83670260	523.469	530.566

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060123\
Data File : P0095422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 04:34
Operator : YP/AJ
Sample : P0060123ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO060123

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 06:54:06 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:53:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

