

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060123\
 Data File : P0095426.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 05:41
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO060123AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 06:30:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:29:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.437	3.655	197.1E6	89859192	54.421	54.927
2) SA Decachlor...	10.278	8.715	233.8E6	116.2E6	50.670	50.180
Target Compounds						
41) L9 AR-1268-1	8.729	7.594	188.0E6	108.1E6	509.888	500.107
42) L9 AR-1268-2	8.824	7.659	171.2E6	96650672	505.461	506.368
43) L9 AR-1268-3	9.060	7.868	147.1E6	82261200	508.684	507.470
44) L9 AR-1268-4	9.496	8.157	65574947	34789720	515.636	505.669
45) L9 AR-1268-5	9.927	8.454	476.0E6	243.6E6	506.067	507.792

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060123\
Data File : PO095426.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 05:41
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO060123AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 06:30:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:29:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

