

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060123\
 Data File : P0095416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 Jun 2023 02:53
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/02/2023
 Supervised By :Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 02 06:33:39 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:33:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.437	3.655	187.2E6	84892715	50.000	50.000
2) SA Decachlor...	10.279	8.715	141.1E6	69052749	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.857	6.858	92292932	23148632	500.000	500.000
37) L8 AR-1262-2	8.411	7.065	159.6E6	48767614	500.000	500.000
38) L8 AR-1262-3	8.732	7.595	111.0E6	38330286	500.000	500.000
39) L8 AR-1262-4	8.822	7.659	58871396	68725363	493.472m	500.000
40) L8 AR-1262-5	9.497	8.157	59977431	31803766	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060123\
Data File : P0095416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 Jun 2023 02:53
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 02 06:33:39 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:33:23 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.437 min
Delta R.T.: 0.000 min
Response: 187247883
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023




#1 Tetrachloro-m-xylene

R.T.: 3.655 min
Delta R.T.: 0.000 min
Response: 84892715
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.279 min
Delta R.T.: 0.000 min
Response: 141110219
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.715 min
Delta R.T.: 0.000 min
Response: 69052749
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 7.857 min
Delta R.T.: 0.000 min
Response: 92292932
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023




#36 AR-1262-1

R.T.: 6.858 min
Delta R.T.: 0.000 min
Response: 23148632
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.411 min
Delta R.T.: 0.000 min
Response: 159637559
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.065 min
Delta R.T.: 0.000 min
Response: 48767614
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 110998958
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023




#38 AR-1262-3

R.T.: 7.595 min
Delta R.T.: 0.000 min
Response: 38330286
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 58871396
Conc: 493.47 ng/ml m



#39 AR-1262-4

R.T.: 7.659 min
Delta R.T.: 0.000 min
Response: 68725363
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 9.497 min
Delta R.T.: 0.000 min
Response: 59977431
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/02/2023
Supervised By :Ankita Jodhani 06/02/2023



#40 AR-1262-5

R.T.: 8.157 min
Delta R.T.: 0.000 min
Response: 31803766
Conc: 500.00 ng/ml