

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060421\
 Data File : P0078438.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 04 Jun 2021 13:06
 Operator : DD\AJ
 Sample : PB136820BL
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB136820BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 05 01:36:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed May 19 06:28:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.345	3.616	1444614	639326	19.546	21.158
2) SA Decachlor...	9.955	8.803	928004	562434	18.770	19.342

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060421\
Data File : P0078438.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 Jun 2021 13:06
Operator : DD\AJ
Sample : PB136820BL
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB136820BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 05 01:36:09 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051821.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed May 19 06:28:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.345 min
Delta R.T.: 0.001 min
Response: 1444614
Conc: 19.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB136820BL




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.008 min
Response: 639326
Conc: 21.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.955 min
Delta R.T.: 0.000 min
Response: 928004
Conc: 18.77 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.803 min
Delta R.T.: -0.013 min
Response: 562434
Conc: 19.34 ng/ml