

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060523\
 Data File : P0095488.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 Jun 2023 22:06
 Operator : YP/AJ
 Sample : 03029-01
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TP-6

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 05:28:40 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:53:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.438	3.654	99754092	45800787	27.066	29.200
2) SA Decachlor...	10.283	8.712	75850851	38859918	28.026	30.038

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060523\
Data File : PO095488.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 Jun 2023 22:06
Operator : YP/AJ
Sample : 03029-01
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-6

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 05:28:40 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:53:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.438 min
Delta R.T.: 0.000 min
Response: 99754092
Conc: 27.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-6




#1 Tetrachloro-m-xylene

R.T.: 3.654 min
Delta R.T.: -0.002 min
Response: 45800787
Conc: 29.20 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.283 min
Delta R.T.: 0.005 min
Response: 75850851
Conc: 28.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.712 min
Delta R.T.: -0.005 min
Response: 38859918
Conc: 30.04 ng/ml