

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP060722.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jun 07 09:47:55 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP048471.D 1000 =PP048469.D 750 =PP048470.D
 250 =PP048472.D 50 =PP048473.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.554	2.394	2.463	2.566	2.414	2.478	E6	3.18
2) SA	Decachlorobiphenyl	2.031	1.878	1.958	2.089	1.901	1.971	E6	4.49
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1	6.833	6.339	6.523	7.051	7.009	6.751	E4	4.60
17) L4	AR-1242-2	0.979	0.900	0.945	1.012	1.023	0.972	E5	5.19
18) L4	AR-1242-3	6.048	5.555	5.808	6.445	6.457	6.063	E4	6.52
19) L4	AR-1242-4	5.061	4.633	4.831	5.308	5.602	5.087	E4	7.53
20) L4	AR-1242-5	5.247	4.822	4.994	5.575	6.062	5.340	E4	9.24
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =PP048471.D 1000 =PP048469.D 750 =PP048470.D
 250 =PP048472.D 50 =PP048473.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	2.012	1.903	1.950	2.056	2.043	1.993	E6	3.25
2) SA	Decachlorobiphenyl	1.242	1.136	1.188	1.295	1.294	1.231	E6	5.61
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\
 Method File : PP060722.M
 Title : GC EXTRACTABLES
 Last Update : Tue Jun 07 09:47:55 2022
 Response Via : Initial Calibration

Calibration Files

500 =PP048471.D 1000 =PP048469.D 750 =PP048470.D
 250 =PP048472.D 50 =PP048473.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	5.570	5.047	5.231	6.007	5.963	5.564 E4	7.69
17) L4	AR-1242-2	7.628	6.956	7.248	8.040	8.089	7.592 E4	6.49
18) L4	AR-1242-3	4.273	3.854	4.031	4.541	4.407	4.221 E4	6.60
19) L4	AR-1242-4	4.088	3.626	3.814	4.345	4.248	4.024 E4	7.46
20) L4	AR-1242-5	4.999	4.462	4.679	5.382	5.173	4.939 E4	7.51
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range