

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060623\
 Data File : P0095517.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2023 14:22
 Operator : YP/AJ
 Sample : 03043-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 MR-PIPE-A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2023
 Supervised By :Ankita Jodhani 06/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 21:06:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:53:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.440	3.656	113.3E6	51304770	30.740	32.709
2) SA Decachlor...	10.294	8.718	71386546	32969848	26.376m	25.485m
Target Compounds						
16) L4 AR-1242-1	5.622	4.751	4001059	3184758	43.229m	71.391 #
17) L4 AR-1242-2	5.644	4.770	6265961	2274464	49.162m	39.147
18) L4 AR-1242-3	5.707	4.948	5079106	1353866	62.413m	42.994 #
19) L4 AR-1242-4	5.813	5.033	5488865	1389440	85.149	41.191 #
20) L4 AR-1242-5	6.551	5.560	2759137	1900078	39.966m	44.315

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060623\
Data File : PO095517.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2023 14:22
Operator : YP/AJ
Sample : 03043-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MR-PIPE-A

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2023
Supervised By :Ankita Jodhani 06/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 21:06:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:53:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

