

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0060623\
 Data File : P0095522.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2023 15:46
 Operator : YP/AJ
 Sample : 03043-06
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 MR-PIPE-F

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2023
 Supervised By :Ankita Jodhani 06/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 21:13:33 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:53:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.440	3.657	95552570	44599510	25.926	28.434
2) SA Decachlor...	10.295	8.718	71256060	35588799	26.328	27.510
Target Compounds						
16) L4 AR-1242-1	5.622	4.751	6694670	6136529	72.332m	137.559 #
17) L4 AR-1242-2	5.645	4.770	11868967	5443317	93.123m	93.689
18) L4 AR-1242-3	5.707	4.947	10208154	4775813	125.439m	151.663
19) L4 AR-1242-4	5.810	5.033	15050490	7796834	233.478	231.141
20) L4 AR-1242-5	6.552	5.560	19343471	9773672	280.191	227.950

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060623\
Data File : PO095522.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2023 15:46
Operator : YP/AJ
Sample : 03043-06
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MR-PIPE-F

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2023
Supervised By :Ankita Jodhani 06/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 21:13:33 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:53:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

