

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0060623\
 Data File : P0095533.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2023 19:35
 Operator : YP/AJ
 Sample : 03043-12
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 MR-PIPE-L

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2023
 Supervised By :Ankita Jodhani 06/07/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 21:24:25 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0060123.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 02 06:53:15 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.438	3.656	84279069	44405652	22.867	28.311
2) SA Decachlor...	10.291	8.714	69557802	31617747	25.701	24.440
Target Compounds						
16) L4 AR-1242-1	5.621	4.749	39914345	41757697	431.254	936.058 #
17) L4 AR-1242-2	5.645	4.768	106.1E6	27340927	832.140	470.584 #
18) L4 AR-1242-3	5.706	4.945	49341746	13617239	606.316	432.435m#
19) L4 AR-1242-4	5.812	5.031	82385176	19927176	1278.037	590.751 #
20) L4 AR-1242-5	6.571	5.558	361.3E6	20388688	5234.082	475.522 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060623\
Data File : PO095533.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2023 19:35
Operator : YP/AJ
Sample : 03043-12
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
MR-PIPE-L

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2023
Supervised By :Ankita Jodhani 06/07/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 21:24:25 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060123.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 02 06:53:15 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

