

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060624\
 Data File : P0104231.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2024 16:18
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 06 16:33:45 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 06 16:32:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.372	3.516	384.1E6	107.3E6	50.000	50.000
2) SA Decachlor...	10.069	8.434	281.2E6	157.3E6	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.535	4.577	126.9E6	33746476	500.000	500.000
4) L1 AR-1016-2	5.557	4.594	177.8E6	51349467	500.000	500.000
5) L1 AR-1016-3	5.619	4.767	110.9E6	28846175	500.000	500.000
6) L1 AR-1016-4	5.718	4.809	91831927	18701909	500.000	500.000
7) L1 AR-1016-5	6.012	5.018	88998102	27348723	500.000	500.000
31) L7 AR-1260-1	7.138	6.038	159.2E6	51713035	500.000	500.000
32) L7 AR-1260-2	7.394	6.225	185.0E6	63484076	500.000	500.000
33) L7 AR-1260-3	7.753	6.377	138.4E6	62773694	500.000	500.000
34) L7 AR-1260-4	7.978	6.845	157.9E6	53253980	500.000	500.000
35) L7 AR-1260-5	8.293	7.086	297.8E6	135.8E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060624\
Data File : PO104231.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 16:18
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 06 16:33:45 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060624.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 06 16:32:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.372 min
Delta R.T.: 0.000 min
Response: 384067528
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.516 min
Delta R.T.: 0.000 min
Response: 107296782
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.069 min
Delta R.T.: 0.000 min
Response: 281171690
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.434 min
Delta R.T.: 0.000 min
Response: 157294365
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.535 min
Delta R.T.: 0.000 min
Response: 126943081
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.577 min
Delta R.T.: 0.000 min
Response: 33746476
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.557 min
Delta R.T.: 0.000 min
Response: 177843205
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.594 min
Delta R.T.: 0.000 min
Response: 51349467
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.619 min
Delta R.T.: 0.000 min
Response: 110859361
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.767 min
Delta R.T.: 0.000 min
Response: 28846175
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.718 min
Delta R.T.: 0.000 min
Response: 91831927
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.809 min
Delta R.T.: 0.000 min
Response: 18701909
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.012 min
Delta R.T.: 0.000 min
Response: 88998102
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.018 min
Delta R.T.: 0.000 min
Response: 27348723
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.138 min
Delta R.T.: 0.000 min
Response: 159216575
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.038 min
Delta R.T.: 0.000 min
Response: 51713035
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.394 min
Delta R.T.: 0.000 min
Response: 185019835
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.225 min
Delta R.T.: 0.000 min
Response: 63484076
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.753 min
Delta R.T.: 0.000 min
Response: 138370320
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.377 min
Delta R.T.: 0.000 min
Response: 62773694
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.978 min
Delta R.T.: 0.000 min
Response: 157932515
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.845 min
Delta R.T.: 0.000 min
Response: 53253980
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.293 min
Delta R.T.: 0.000 min
Response: 297824006
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.086 min
Delta R.T.: 0.000 min
Response: 135819851
Conc: 500.00 ng/ml