

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060624\
 Data File : P0104246.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2024 20:33
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 01:01:26 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 07 00:59:53 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.371	3.516	858.8E6	256.3E6	93.110	94.437
2) SA Decachlor...	10.069	8.435	557.0E6	315.2E6	91.714	91.216
Target Compounds						
26) L6 AR-1254-1	6.389	5.366	283.1E6	113.0E6	873.338	884.557
27) L6 AR-1254-2	6.607	5.511	414.1E6	97547479	873.795	875.125
28) L6 AR-1254-3	6.973	5.910	438.6E6	174.0E6	888.477	879.782
29) L6 AR-1254-4	7.258	6.136	321.7E6	109.8E6	891.289	890.848
30) L6 AR-1254-5	7.677	6.549	339.6E6	159.1E6	914.017	896.314



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060624\
Data File : PO104246.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 20:33
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 01:01:26 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 07 00:59:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.371 min
Delta R.T.: 0.000 min
Response: 858790301
Conc: 93.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.516 min
Delta R.T.: 0.000 min
Response: 256291408
Conc: 94.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.069 min
Delta R.T.: 0.002 min
Response: 557021068
Conc: 91.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.435 min
Delta R.T.: 0.000 min
Response: 315202956
Conc: 91.22 ng/ml



#26 AR-1254-1


R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 283147643
Conc: 873.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.366 min
Delta R.T.: 0.000 min
Response: 113042011
Conc: 884.56 ng/ml




#27 AR-1254-2

R.T.: 6.607 min
Delta R.T.: 0.000 min
Response: 414104383
Conc: 873.79 ng/ml



#27 AR-1254-2


R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 97547479
Conc: 875.13 ng/ml



#28 AR-1254-3


R.T.: 6.973 min
Delta R.T.: 0.001 min
Response: 438610594
Conc: 888.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 5.910 min
Delta R.T.: 0.000 min
Response: 173992260
Conc: 879.78 ng/ml




#29 AR-1254-4

R.T.: 7.258 min
Delta R.T.: 0.000 min
Response: 321665085
Conc: 891.29 ng/ml



#29 AR-1254-4


R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 109830433
Conc: 890.85 ng/ml



#30 AR-1254-5

R.T.: 7.677 min
Delta R.T.: 0.000 min
Response: 339588738
Conc: 914.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.549 min
Delta R.T.: 0.000 min
Response: 159069420
Conc: 896.31 ng/ml