

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060624\
 Data File : P0104249.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 06 Jun 2024 21:24
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/07/2024
 Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 07 01:08:19 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 07 00:59:53 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.371	3.516	233.5E6	67063801	25.770	25.173
2)	SA Decachlor...	10.068	8.434	160.3E6	90953970	26.831	26.846
Target Compounds							
26)	L6 AR-1254-1	6.389	5.365	86768471	34062559	274.423	273.006
27)	L6 AR-1254-2	6.606	5.511	127.4E6	30169731	276.541m	277.019
28)	L6 AR-1254-3	6.972	5.909	129.9E6	51340118	269.779	269.401
29)	L6 AR-1254-4	7.258	6.136	94852389	32348166	269.390	269.451
30)	L6 AR-1254-5	7.676	6.550	98915188	46530394	269.966	268.829

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060624\
Data File : PO104249.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 06 Jun 2024 21:24
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 07 01:08:19 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060624.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 07 00:59:53 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.371 min
Delta R.T.: 0.000 min
Response: 233461956
Conc: 25.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024




#1 Tetrachloro-m-xylene

R.T.: 3.516 min
Delta R.T.: 0.000 min
Response: 67063801
Conc: 25.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.068 min
Delta R.T.: 0.000 min
Response: 160344611
Conc: 26.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.434 min
Delta R.T.: 0.000 min
Response: 90953970
Conc: 26.85 ng/ml




#26 AR-1254-1

R.T.: 6.389 min
Delta R.T.: 0.000 min
Response: 86768471
Conc: 274.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254\ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024




#26 AR-1254-1

R.T.: 5.365 min
Delta R.T.: 0.000 min
Response: 34062559
Conc: 273.01 ng/ml




#27 AR-1254-2

R.T.: 6.606 min
Delta R.T.: 0.000 min
Response: 127360279
Conc: 276.54 ng/ml m



#27 AR-1254-2

R.T.: 5.511 min
Delta R.T.: 0.000 min
Response: 30169731
Conc: 277.02 ng/ml




#28 AR-1254-3

R.T.: 6.972 min
Delta R.T.: 0.000 min
Response: 129939892
Conc: 269.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024




#28 AR-1254-3

R.T.: 5.909 min
Delta R.T.: 0.000 min
Response: 51340118
Conc: 269.40 ng/ml




#29 AR-1254-4

R.T.: 7.258 min
Delta R.T.: 0.000 min
Response: 94852389
Conc: 269.39 ng/ml



#29 AR-1254-4

R.T.: 6.136 min
Delta R.T.: 0.000 min
Response: 32348166
Conc: 269.45 ng/ml




#30 AR-1254-5

R.T.: 7.676 min
Delta R.T.: 0.000 min
Response: 98915188
Conc: 269.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/07/2024
Supervised By :Ankita Jodhani 06/07/2024



#30 AR-1254-5

R.T.: 6.550 min
Delta R.T.: 0.000 min
Response: 46530394
Conc: 268.83 ng/ml