

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060721\
 Data File : P0078545.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2021 2:43
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO060721AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 05:45:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	76.380	73.406 E3	3.9	94	0.00
2 SA Decachlorobiphenyl	40.548	71.867 E3	-77.2#	177#	0.00
41 L9 AR-1268-1	7.173	6.954 E3	3.1	99	0.00
42 L9 AR-1268-2	6.371	6.234 E3	2.2	100	0.00
43 L9 AR-1268-3	5.194	5.316 E3	-2.3	104	0.00
44 L9 AR-1268-4	1.980	1.969 E3	0.6	99	0.00
45 L9 AR-1268-5	15.386	15.066 E3	2.1	99	0.00

Signal #2

1 SA Tetrachloro-m-xylene	31.593	31.173 E3	1.3	94	0.00
2 SA Decachlorobiphenyl	24.045	45.945 E3	-91.1#	191#	0.00
41 L9 AR-1268-1	4.542	4.195 E3	7.6	96	0.00
42 L9 AR-1268-2	4.095	3.800 E3	7.2	96	0.00
43 L9 AR-1268-3	3.489	3.283 E3	5.9	97	0.00
44 L9 AR-1268-4	1.468	1.415 E3	3.6	98	0.00
45 L9 AR-1268-5	10.358	9.769 E3	5.7	97	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	2.342	0.000 E3	100.0#	0#	-5.64#
4 L1 AR-1016-2	3.605	0.000 E3	100.0#	0#	-5.66#
5 L1 AR-1016-3	2.187	0.000 E3	100.0#	0#	-5.72#
6 L1 AR-1016-4	1.754	0.000 E3	100.0#	0#	-5.83#
7 L1 AR-1016-5	1.735	0.000 E3	100.0#	0#	-6.14#
8 L2 AR-1221-1	836.422	0.000	100.0#	0#	-4.59#
9 L2 AR-1221-2	605.681	0.000	100.0#	0#	-4.68#
10 L2 AR-1221-3	1.982	0.000 E3	100.0#	0#	-4.77#
11 L3 AR-1232-1	1.531	0.000 E3	100.0#	0#	-4.77#
12 L3 AR-1232-2	777.377	0.000	100.0#	0#	-5.34#
13 L3 AR-1232-3	1.458	0.000 E3	100.0#	0#	-5.66#
14 L3 AR-1232-4	701.624	0.000	100.0#	0#	-5.83#
15 L3 AR-1232-5	503.739	0.000	100.0#	0#	-5.92#
16 L4 AR-1242-1	1.806	0.000 E3	100.0#	0#	-5.63#
17 L4 AR-1242-2	2.683	0.000 E3	100.0#	0#	-5.66#
18 L4 AR-1242-3	1.646	0.000 E3	100.0#	0#	-5.72#
19 L4 AR-1242-4	1.347	0.000 E3	100.0#	0#	-5.82#
20 L4 AR-1242-5	1.338	0.000 E3	100.0#	0#	-6.59#
21 L5 AR-1248-1	1.360	0.000 E3	100.0#	0#	-5.63#
22 L5 AR-1248-2	1.942	0.000 E3	100.0#	0#	-5.92#
23 L5 AR-1248-3	2.260	0.000 E3	100.0#	0#	-6.13#
24 L5 AR-1248-4	2.369	0.000 E3	100.0#	0#	-6.56#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060721\
 Data File : P0078545.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2021 2:43
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO060721AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 05:45:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
25 L5 AR-1248-5	2.327	0.000 E3	100.0#	0#	-6.60#
26 L6 AR-1254-1	2.531	0.000 E3	100.0#	0#	-6.53#
27 L6 AR-1254-2	3.971	0.000 E3	100.0#	0#	-6.75#
28 L6 AR-1254-3	4.154	0.000 E3	100.0#	0#	-7.13#
29 L6 AR-1254-4	2.940	0.000 E3	100.0#	0#	-7.42#
30 L6 AR-1254-5	3.206	0.000 E3	100.0#	0#	-7.85#
31 L7 AR-1260-1	2.932	0.000 E3	100.0#	0#	-7.30#
32 L7 AR-1260-2	3.290	0.000 E3	100.0#	0#	-7.56#
33 L7 AR-1260-3	2.368	0.000 E3	100.0#	0#	-7.92#
34 L7 AR-1260-4	2.572	0.000 E3	100.0#	0#	-8.15#
35 L7 AR-1260-5	5.085	0.000 E3	100.0#	0#	-8.46#
36 L8 AR-1262-1	3.912	0.000 E3	100.0#	0#	-7.92#
37 L8 AR-1262-2	6.398	0.000 E3	100.0#	0#	-8.46#
38 L8 AR-1262-3	4.177	0.000 E3	100.0#	0#	-8.73#
39 L8 AR-1262-4	3.099	0.000 E3	100.0#	0#	-8.81#
40 L8 AR-1262-5	1.895	0.000 E3	100.0#	0#	-9.35#

Signal #2

3 L1 AR-1016-1	1.046	0.000 E3	100.0#	0#	-4.84#
4 L1 AR-1016-2	1.579	0.000 E3	100.0#	0#	-4.86#
5 L1 AR-1016-3	800.774	0.000	100.0#	0#	-5.05#
6 L1 AR-1016-4	706.202	0.000	100.0#	0#	-5.10#
7 L1 AR-1016-5	872.212	0.000	100.0#	0#	-5.33#
8 L2 AR-1221-1	380.786	0.000	100.0#	0#	-3.87#
9 L2 AR-1221-2	281.230	0.000	100.0#	0#	-3.96#
10 L2 AR-1221-3	900.768	0.000	100.0#	0#	-4.05#
11 L3 AR-1232-1	727.021	0.000	100.0#	0#	-4.05#
12 L3 AR-1232-2	688.994	0.000	100.0#	0#	-4.86#
13 L3 AR-1232-3	342.455	0.000	100.0#	0#	-5.05#
14 L3 AR-1232-4	307.509	0.000	100.0#	0#	-5.14#
15 L3 AR-1232-5	348.595	0.000	100.0#	0#	-5.33#
16 L4 AR-1242-1	841.077	0.000	100.0#	0#	-4.84#
17 L4 AR-1242-2	1.230	0.000 E3	100.0#	0#	-4.86#
18 L4 AR-1242-3	634.161	0.000	100.0#	0#	-5.05#
19 L4 AR-1242-4	654.298	0.000	100.0#	0#	-5.14#
20 L4 AR-1242-5	822.230	0.000	100.0#	0#	-5.70#
21 L5 AR-1248-1	646.276	0.000	100.0#	0#	-4.84#
22 L5 AR-1248-2	975.070	0.000	100.0#	0#	-5.10#
23 L5 AR-1248-3	970.209	0.000	100.0#	0#	-5.14#
24 L5 AR-1248-4	1.158	0.000 E3	100.0#	0#	-5.32#
25 L5 AR-1248-5	1.096	0.000 E3	100.0#	0#	-5.74#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060721\
 Data File : P0078545.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2021 2:43
 Operator : DD\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO060721AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 05:45:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
26 L6 AR-1254-1	1.726	0.000 E3	100.0#	0#	-5.70#
27 L6 AR-1254-2	1.603	0.000 E3	100.0#	0#	-5.86#
28 L6 AR-1254-3	2.393	0.000 E3	100.0#	0#	-6.27#
29 L6 AR-1254-4	1.463	0.000 E3	100.0#	0#	-6.51#
30 L6 AR-1254-5	2.280	0.000 E3	100.0#	0#	-6.93#
31 L7 AR-1260-1	1.587	0.000 E3	100.0#	0#	-6.41#
32 L7 AR-1260-2	1.863	0.000 E3	100.0#	0#	-6.60#
33 L7 AR-1260-3	1.774	0.000 E3	100.0#	0#	-6.75#
34 L7 AR-1260-4	1.465	0.000 E3	100.0#	0#	-7.23#
35 L7 AR-1260-5	3.489	0.000 E3	100.0#	0#	-7.48#
36 L8 AR-1262-1	1.195	0.000 E3	100.0#	0#	-6.93#
37 L8 AR-1262-2	3.922	0.000 E3	100.0#	0#	-7.48#
38 L8 AR-1262-3	1.646	0.000 E3	100.0#	0#	-7.76#
39 L8 AR-1262-4	3.008	0.000 E3	100.0#	0#	-7.82#
40 L8 AR-1262-5	1.357	0.000 E3	100.0#	0#	-8.31#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0