

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060721\
 Data File : P0078522.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jun 2021 20:06
 Operator : DD\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 03:52:40 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 03:51:05 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.337	3.613	358029	168959	4.734	5.155
2)	SA Decachlor...	9.947	8.799	233861	156545	5.602	6.024
Target Compounds							
21)	L5 AR-1248-1	5.633	4.842	67232	34751	48.526	53.654
22)	L5 AR-1248-2	5.924	5.101	96487	55795	50.634	58.261
23)	L5 AR-1248-3	6.134	5.144	119192	54745	52.839	57.435
24)	L5 AR-1248-4	6.558	5.324	124201	65940	52.313	57.893
25)	L5 AR-1248-5	6.596	5.740	118477	48724	51.191	45.655



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060721\
Data File : PO078522.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jun 2021 20:06
Operator : DD\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 03:52:40 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 03:51:05 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.337 min
Delta R.T.: 0.000 min
Response: 358029
Conc: 4.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.613 min
Delta R.T.: 0.000 min
Response: 168959
Conc: 5.15 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.947 min
Delta R.T.: 0.000 min
Response: 233861
Conc: 5.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.799 min
Delta R.T.: 0.000 min
Response: 156545
Conc: 6.02 ng/ml



#21 AR-1248-1


R.T.: 5.633 min
Delta R.T.: 0.000 min
Response: 67232
Conc: 48.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.842 min
Delta R.T.: 0.000 min
Response: 34751
Conc: 53.65 ng/ml




#22 AR-1248-2

R.T.: 5.924 min
Delta R.T.: 0.000 min
Response: 96487
Conc: 50.63 ng/ml



#22 AR-1248-2


R.T.: 5.101 min
Delta R.T.: 0.000 min
Response: 55795
Conc: 58.26 ng/ml



#23 AR-1248-3


R.T.: 6.134 min
Delta R.T.: 0.000 min
Response: 119192
Conc: 52.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 5.144 min
Delta R.T.: 0.000 min
Response: 54745
Conc: 57.44 ng/ml




#24 AR-1248-4

R.T.: 6.558 min
Delta R.T.: 0.000 min
Response: 124201
Conc: 52.31 ng/ml



#24 AR-1248-4


R.T.: 5.324 min
Delta R.T.: 0.000 min
Response: 65940
Conc: 57.89 ng/ml



#25 AR-1248-5

R.T.: 6.596 min
Delta R.T.: 0.000 min
Response: 118477
Conc: 51.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.740 min
Delta R.T.: 0.000 min
Response: 48724
Conc: 45.66 ng/ml