

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060721\
 Data File : P0078540.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2021 1:17
 Operator : DD\AJ
 Sample : AR1232ICV500
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO060721AR1232

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 05:42:39 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.338	3.614	3754075	1513495	49.150	47.906
2) SA Decachlor...	9.946	8.799	2099499	1366812	51.778	56.843
Target Compounds						
11) L3 AR-1232-1	4.768	4.049	770996	345275	503.600	474.918
12) L3 AR-1232-2	5.343	4.861	400600	317899	515.322	461.396
13) L3 AR-1232-3	5.656	5.048	703290	164635	482.310	480.750
14) L3 AR-1232-4	5.825	5.144	347649	154273	495.492	501.686
15) L3 AR-1232-5	5.926	5.325	307056	167739	609.554	481.186



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060721\
Data File : P0078540.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jun 2021 1:17
Operator : DD\AJ
Sample : AR1232ICV500
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO060721AR1232

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 05:42:39 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 05:39:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.338 min
Delta R.T.: -0.001 min
Response: 3754075
Conc: 49.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO060721AR1232




#1 Tetrachloro-m-xylene

R.T.: 3.614 min
Delta R.T.: -0.001 min
Response: 1513495
Conc: 47.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.946 min
Delta R.T.: -0.004 min
Response: 2099499
Conc: 51.78 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.799 min
Delta R.T.: -0.003 min
Response: 1366812
Conc: 56.84 ng/ml



#11 AR-1232-1


R.T.: 4.768 min
Delta R.T.: 0.000 min
Response: 770996
Conc: 503.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO060721AR1232




#11 AR-1232-1

R.T.: 4.049 min
Delta R.T.: 0.000 min
Response: 345275
Conc: 474.92 ng/ml




#12 AR-1232-2

R.T.: 5.343 min
Delta R.T.: 0.000 min
Response: 400600
Conc: 515.32 ng/ml



#12 AR-1232-2


R.T.: 4.861 min
Delta R.T.: 0.000 min
Response: 317899
Conc: 461.40 ng/ml



#13 AR-1232-3


R.T.: 5.656 min
Delta R.T.: 0.000 min
Response: 703290
Conc: 482.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO060721AR1232




#13 AR-1232-3

R.T.: 5.048 min
Delta R.T.: 0.000 min
Response: 164635
Conc: 480.75 ng/ml




#14 AR-1232-4

R.T.: 5.825 min
Delta R.T.: 0.000 min
Response: 347649
Conc: 495.49 ng/ml



#14 AR-1232-4


R.T.: 5.144 min
Delta R.T.: 0.000 min
Response: 154273
Conc: 501.69 ng/ml



#15 AR-1232-5

R.T.: 5.926 min
Delta R.T.: 0.001 min
Response: 307056
Conc: 609.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO060721AR1232



#15 AR-1232-5

R.T.: 5.325 min
Delta R.T.: -0.001 min
Response: 167739
Conc: 481.19 ng/ml