

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060722\
 Data File : P0086989.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jun 2022 18:02
 Operator : YP\AJ
 Sample : N3143-05MS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 01:32:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 17:55:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.513	3.660	1803567	847246	31.536	26.506
2)	SA Decachlor...	10.387	8.683	1026740	438100	22.494	19.538
Target Compounds							
3)	L1 AR-1016-1	5.698	4.751	1525086	929717	768.053	941.857
4)	L1 AR-1016-2	5.721	4.772	2146967	827496	757.655	606.744
5)	L1 AR-1016-3	5.784	4.947	1361988	452398	779.053	613.702
6)	L1 AR-1016-4	5.883	4.991	1080771	348313	773.770	596.338
7)	L1 AR-1016-5	6.181	5.204	1014842	446092	738.577	612.668
8)	L2 AR-1221-1	4.717	3.876	154185	64113	232.090	167.960 #
9)	L2 AR-1221-2	4.812	3.963	205777	80653	392.672	284.676 #
10)	L2 AR-1221-3	4.889	4.039	808323	357468	501.071	424.419
11)	L3 AR-1232-1	4.889	4.039	808323	357468	529.348	441.690
12)	L3 AR-1232-2	5.231	4.772	1149681	827496	1635.790	1249.114
13)	L3 AR-1232-3	5.721	4.947	2146967	452398	1583.397	1312.488
14)	L3 AR-1232-4	5.883	5.033	1080771	430528	1654.652	1420.362
15)	L3 AR-1232-5	5.975	5.204	952081	446092	1817.112	1350.031 #
16)	L4 AR-1242-1	5.698	4.751	1525086	929717	954.072	1183.471
17)	L4 AR-1242-2	5.721	4.772	2146967	827496	942.396	765.574
18)	L4 AR-1242-3	5.784	4.947	1361988	452398	958.969	766.257
19)	L4 AR-1242-4	5.883	5.033	1080771	430528	953.608	752.375
20)	L4 AR-1242-5	6.627	5.557	138680	350490	117.932	521.883 #
21)	L5 AR-1248-1	5.698	4.751	1525086	929717	1250.621	1581.227 #
22)	L5 AR-1248-2	5.975	4.991	952081	348313	544.294	432.667
23)	L5 AR-1248-3	6.181	5.033	1014842	430528	525.094	513.105
24)	L5 AR-1248-4	6.562	5.204	908686	446092	423.078	458.550
25)	L5 AR-1248-5	6.627	5.597	138680	52808	67.657	56.808
26)	L6 AR-1254-1	6.562	5.557	908686	350490	408.960	243.706 #
27)	L6 AR-1254-2	6.781	5.704	796300	322526	232.422	260.323
28)	L6 AR-1254-3	7.151	6.124	399792	582617	114.433	290.040 #
29)	L6 AR-1254-4	7.439	6.336	245689	63916	95.782	51.794 #
30)	L6 AR-1254-5	7.859	6.754	2272164	1072866	821.483	619.395
31)	L7 AR-1260-1	7.317	6.239	1668879	802899	672.447	620.007
32)	L7 AR-1260-2	7.574	6.426	1974398	940637	582.386	608.990
33)	L7 AR-1260-3	7.937	6.579	1206877	928195	558.540	534.100
34)	L7 AR-1260-4	8.166	7.052	1472250	600734	572.937	527.796
35)	L7 AR-1260-5	8.495	7.293	2618446	1364558	551.032	513.801
36)	L8 AR-1262-1	7.937	6.792	1206877	643617	363.375	372.213
37)	L8 AR-1262-2	8.495	7.293	2618446	1364558	457.638	454.394
38)	L8 AR-1262-3	8.834	7.577	1701594	247065	432.141	206.077 #
39)	L8 AR-1262-4	8.894	7.639	955117	949232	318.919	434.296 #
40)	L8 AR-1262-5	9.597	8.134	612347	109774	296.140	108.838 #
41)	L9 AR-1268-1	8.834	7.577	1701594	247065	242.252	72.096 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060722\
 Data File : P0086989.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jun 2022 18:02
 Operator : YP\AJ
 Sample : N3143-05MS
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 01:32:54 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 17:55:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.894	7.639	955117	949232	149.174	302.682 #
43) L9	AR-1268-3	9.153	7.850	40683	20807	7.480	7.862
44) L9	AR-1268-4	9.597	8.134	612347	109774	259.308	96.097 #
45) L9	AR-1268-5	10.030	8.427	192714	75698	10.711	8.667



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060722\
Data File : PO086989.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jun 2022 18:02
Operator : YP\AJ
Sample : N3143-05MS
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 01:32:54 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 17:55:39 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 1803567
Conc: 31.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.660 min
Delta R.T.: -0.006 min
Response: 847246
Conc: 26.51 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.387 min
Delta R.T.: 0.003 min
Response: 1026740
Conc: 22.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: -0.018 min
Response: 438100
Conc: 19.54 ng/ml



#3 AR-1016-1


R.T.: 5.698 min
Delta R.T.: 0.000 min
Response: 1525086
Conc: 768.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.751 min
Delta R.T.: -0.010 min
Response: 929717
Conc: 941.86 ng/ml




#4 AR-1016-2

R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 2146967
Conc: 757.66 ng/ml



#4 AR-1016-2


R.T.: 4.772 min
Delta R.T.: -0.008 min
Response: 827496
Conc: 606.74 ng/ml



#5 AR-1016-3


R.T.: 5.784 min
Delta R.T.: 0.001 min
Response: 1361988
Conc: 779.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.947 min
Delta R.T.: -0.009 min
Response: 452398
Conc: 613.70 ng/ml




#6 AR-1016-4

R.T.: 5.883 min
Delta R.T.: -0.001 min
Response: 1080771
Conc: 773.77 ng/ml



#6 AR-1016-4


R.T.: 4.991 min
Delta R.T.: -0.009 min
Response: 348313
Conc: 596.34 ng/ml



#7 AR-1016-5


R.T.: 6.181 min
Delta R.T.: 0.000 min
Response: 1014842
Conc: 738.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 446092
Conc: 612.67 ng/ml




#8 AR-1221-1

R.T.: 4.717 min
Delta R.T.: -0.009 min
Response: 154185
Conc: 232.09 ng/ml



#8 AR-1221-1


R.T.: 3.876 min
Delta R.T.: -0.009 min
Response: 64113
Conc: 167.96 ng/ml



#9 AR-1221-2


R.T.: 4.812 min
Delta R.T.: -0.001 min
Response: 205777
Conc: 392.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.963 min
Delta R.T.: -0.008 min
Response: 80653
Conc: 284.68 ng/ml




#10 AR-1221-3

R.T.: 4.889 min
Delta R.T.: -0.002 min
Response: 808323
Conc: 501.07 ng/ml



#10 AR-1221-3


R.T.: 4.039 min
Delta R.T.: -0.009 min
Response: 357468
Conc: 424.42 ng/ml



#11 AR-1232-1


R.T.: 4.889 min
Delta R.T.: -0.002 min
Response: 808323
Conc: 529.35 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.039 min
Delta R.T.: -0.008 min
Response: 357468
Conc: 441.69 ng/ml




#12 AR-1232-2

R.T.: 5.231 min
Delta R.T.: -0.001 min
Response: 1149681
Conc: 1635.79 ng/ml



#12 AR-1232-2


R.T.: 4.772 min
Delta R.T.: -0.009 min
Response: 827496
Conc: 1249.11 ng/ml



#13 AR-1232-3


R.T.: 5.721 min
Delta R.T.: 0.000 min
Response: 2146967
Conc: 1583.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.947 min
Delta R.T.: -0.010 min
Response: 452398
Conc: 1312.49 ng/ml




#14 AR-1232-4

R.T.: 5.883 min
Delta R.T.: 0.000 min
Response: 1080771
Conc: 1654.65 ng/ml



#14 AR-1232-4


R.T.: 5.033 min
Delta R.T.: -0.009 min
Response: 430528
Conc: 1420.36 ng/ml



#15 AR-1232-5


R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 952081
Conc: 1817.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.204 min
Delta R.T.: -0.010 min
Response: 446092
Conc: 1350.03 ng/ml




#16 AR-1242-1

R.T.: 5.698 min
Delta R.T.: -0.001 min
Response: 1525086
Conc: 954.07 ng/ml



#16 AR-1242-1


R.T.: 4.751 min
Delta R.T.: -0.010 min
Response: 929717
Conc: 1183.47 ng/ml



#17 AR-1242-2


R.T.: 5.721 min
Delta R.T.: -0.001 min
Response: 2146967
Conc: 942.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.772 min
Delta R.T.: -0.009 min
Response: 827496
Conc: 765.57 ng/ml




#18 AR-1242-3

R.T.: 5.784 min
Delta R.T.: -0.002 min
Response: 1361988
Conc: 958.97 ng/ml



#18 AR-1242-3


R.T.: 4.947 min
Delta R.T.: -0.009 min
Response: 452398
Conc: 766.26 ng/ml



#19 AR-1242-4


R.T.: 5.883 min
Delta R.T.: -0.002 min
Response: 1080771
Conc: 953.61 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.033 min
Delta R.T.: -0.008 min
Response: 430528
Conc: 752.37 ng/ml




#20 AR-1242-5

R.T.: 6.627 min
Delta R.T.: 0.000 min
Response: 138680
Conc: 117.93 ng/ml



#20 AR-1242-5


R.T.: 5.557 min
Delta R.T.: -0.008 min
Response: 350490
Conc: 521.88 ng/ml



#21 AR-1248-1


R.T.: 5.698 min
Delta R.T.: -0.001 min
Response: 1525086
Conc: 1250.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.751 min
Delta R.T.: -0.010 min
Response: 929717
Conc: 1581.23 ng/ml




#22 AR-1248-2

R.T.: 5.975 min
Delta R.T.: 0.000 min
Response: 952081
Conc: 544.29 ng/ml



#22 AR-1248-2


R.T.: 4.991 min
Delta R.T.: -0.009 min
Response: 348313
Conc: 432.67 ng/ml



#23 AR-1248-3


R.T.: 6.181 min
Delta R.T.: -0.001 min
Response: 1014842
Conc: 525.09 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.033 min
Delta R.T.: -0.008 min
Response: 430528
Conc: 513.11 ng/ml




#24 AR-1248-4

R.T.: 6.562 min
Delta R.T.: -0.027 min
Response: 908686
Conc: 423.08 ng/ml



#24 AR-1248-4


R.T.: 5.204 min
Delta R.T.: -0.009 min
Response: 446092
Conc: 458.55 ng/ml



#25 AR-1248-5


R.T.: 6.627 min
Delta R.T.: 0.000 min
Response: 138680
Conc: 67.66 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.597 min
Delta R.T.: -0.009 min
Response: 52808
Conc: 56.81 ng/ml




#26 AR-1254-1

R.T.: 6.562 min
Delta R.T.: -0.002 min
Response: 908686
Conc: 408.96 ng/ml



#26 AR-1254-1


R.T.: 5.557 min
Delta R.T.: -0.010 min
Response: 350490
Conc: 243.71 ng/ml



#27 AR-1254-2


R.T.: 6.781 min
Delta R.T.: -0.004 min
Response: 796300
Conc: 232.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.704 min
Delta R.T.: -0.010 min
Response: 322526
Conc: 260.32 ng/ml




#28 AR-1254-3

R.T.: 7.151 min
Delta R.T.: -0.003 min
Response: 399792
Conc: 114.43 ng/ml



#28 AR-1254-3


R.T.: 6.124 min
Delta R.T.: 0.006 min
Response: 582617
Conc: 290.04 ng/ml



#29 AR-1254-4


R.T.: 7.439 min
Delta R.T.: -0.002 min
Response: 245689
Conc: 95.78 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.336 min
Delta R.T.: -0.011 min
Response: 63916
Conc: 51.79 ng/ml




#30 AR-1254-5

R.T.: 7.859 min
Delta R.T.: -0.004 min
Response: 2272164
Conc: 821.48 ng/ml



#30 AR-1254-5


R.T.: 6.754 min
Delta R.T.: -0.012 min
Response: 1072866
Conc: 619.39 ng/ml



#31 AR-1260-1


R.T.: 7.317 min
Delta R.T.: 0.000 min
Response: 1668879
Conc: 672.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.239 min
Delta R.T.: -0.010 min
Response: 802899
Conc: 620.01 ng/ml




#32 AR-1260-2

R.T.: 7.574 min
Delta R.T.: -0.004 min
Response: 1974398
Conc: 582.39 ng/ml



#32 AR-1260-2


R.T.: 6.426 min
Delta R.T.: -0.011 min
Response: 940637
Conc: 608.99 ng/ml



#33 AR-1260-3


R.T.: 7.937 min
Delta R.T.: 0.000 min
Response: 1206877
Conc: 558.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.579 min
Delta R.T.: -0.009 min
Response: 928195
Conc: 534.10 ng/ml




#34 AR-1260-4

R.T.: 8.166 min
Delta R.T.: 0.000 min
Response: 1472250
Conc: 572.94 ng/ml



#34 AR-1260-4


R.T.: 7.052 min
Delta R.T.: -0.012 min
Response: 600734
Conc: 527.80 ng/ml



#35 AR-1260-5


R.T.: 8.495 min
Delta R.T.: -0.001 min
Response: 2618446
Conc: 551.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.293 min
Delta R.T.: -0.012 min
Response: 1364558
Conc: 513.80 ng/ml




#36 AR-1262-1

R.T.: 7.937 min
Delta R.T.: -0.002 min
Response: 1206877
Conc: 363.38 ng/ml



#36 AR-1262-1


R.T.: 6.792 min
Delta R.T.: -0.013 min
Response: 643617
Conc: 372.21 ng/ml



#37 AR-1262-2


R.T.: 8.495 min
Delta R.T.: -0.002 min
Response: 2618446
Conc: 457.64 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.293 min
Delta R.T.: -0.012 min
Response: 1364558
Conc: 454.39 ng/ml




#38 AR-1262-3

R.T.: 8.834 min
Delta R.T.: 0.012 min
Response: 1701594
Conc: 432.14 ng/ml



#38 AR-1262-3


R.T.: 7.577 min
Delta R.T.: -0.013 min
Response: 247065
Conc: 206.08 ng/ml



#39 AR-1262-4


R.T.: 8.894 min
Delta R.T.: -0.020 min
Response: 955117
Conc: 318.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.639 min
Delta R.T.: -0.014 min
Response: 949232
Conc: 434.30 ng/ml




#40 AR-1262-5

R.T.: 9.597 min
Delta R.T.: 0.000 min
Response: 612347
Conc: 296.14 ng/ml



#40 AR-1262-5


R.T.: 8.134 min
Delta R.T.: -0.014 min
Response: 109774
Conc: 108.84 ng/ml



#41 AR-1268-1


R.T.: 8.834 min
Delta R.T.: 0.012 min
Response: 1701594
Conc: 242.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.577 min
Delta R.T.: -0.014 min
Response: 247065
Conc: 72.10 ng/ml




#42 AR-1268-2

R.T.: 8.894 min
Delta R.T.: -0.025 min
Response: 955117
Conc: 149.17 ng/ml



#42 AR-1268-2


R.T.: 7.639 min
Delta R.T.: -0.015 min
Response: 949232
Conc: 302.68 ng/ml



#43 AR-1268-3


R.T.: 9.153 min
Delta R.T.: -0.006 min
Response: 40683
Conc: 7.48 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.850 min
Delta R.T.: -0.013 min
Response: 20807
Conc: 7.86 ng/ml




#44 AR-1268-4

R.T.: 9.597 min
Delta R.T.: -0.002 min
Response: 612347
Conc: 259.31 ng/ml



#44 AR-1268-4


R.T.: 8.134 min
Delta R.T.: -0.014 min
Response: 109774
Conc: 96.10 ng/ml



#45 AR-1268-5

R.T.: 10.030 min
Delta R.T.: -0.006 min
Response: 192714
Conc: 10.71 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.427 min
Delta R.T.: -0.015 min
Response: 75698
Conc: 8.67 ng/ml