

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0060722\
 Data File : P0086999.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 07 Jun 2022 21:29
 Operator : YP\AJ
 Sample : N3181-04 10X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 TRE-31702

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/08/2022
 Supervised By :Ankita Jodhani 06/08/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 08 01:39:51 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0060122.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 01 17:55:39 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.670	97481	58696	1.704m	1.836m
2) SA Decachlor...	10.373	8.685	111800	45687	2.449	2.038m
Target Compounds						
26) L6 AR-1254-1	6.587	5.575	902940	235018	406.374m	163.415m#
27) L6 AR-1254-2	6.794	5.724	439098	248898	128.163	200.895m#
28) L6 AR-1254-3	7.159	6.125	257997	232732	73.847m	115.860m#
29) L6 AR-1254-4	7.426	6.339	362045	452204	141.143m	366.440 #
30) L6 AR-1254-5	7.868	6.769	181663	172431	65.679m	99.549 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO060722\
Data File : PO086999.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 07 Jun 2022 21:29
Operator : YP\AJ
Sample : N3181-04 10X
Misc :
ALS Vial : 34 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
TRE-31702


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/08/2022
Supervised By :Ankita Jodhani 06/08/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 08 01:39:51 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060122.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 01 17:55:39 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.517 min
Delta R.T.: 0.003 min
Response: 97481
Conc: 1.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
TRE-31702


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/08/2022
Supervised By :Ankita Jodhani 06/08/2022




#1 Tetrachloro-m-xylene

R.T.: 3.670 min
Delta R.T.: 0.004 min
Response: 58696
Conc: 1.84 ng/ml m




#2 Decachlorobiphenyl

R.T.: 10.373 min
Delta R.T.: -0.010 min
Response: 111800
Conc: 2.45 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.685 min
Delta R.T.: -0.015 min
Response: 45687
Conc: 2.04 ng/ml m




#26 AR-1254-1

R.T.: 6.587 min
Delta R.T.: 0.023 min
Response: 902940
Conc: 406.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
TRE-31702


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/08/2022
Supervised By :Ankita Jodhani 06/08/2022




#26 AR-1254-1

R.T.: 5.575 min
Delta R.T.: 0.009 min
Response: 235018
Conc: 163.41 ng/ml m




#27 AR-1254-2

R.T.: 6.794 min
Delta R.T.: 0.010 min
Response: 439098
Conc: 128.16 ng/ml



#27 AR-1254-2

R.T.: 5.724 min
Delta R.T.: 0.010 min
Response: 248898
Conc: 200.90 ng/ml m




#28 AR-1254-3

R.T.: 7.159 min
Delta R.T.: 0.006 min
Response: 257997
Conc: 73.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
TRE-31702


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/08/2022
Supervised By :Ankita Jodhani 06/08/2022




#28 AR-1254-3

R.T.: 6.125 min
Delta R.T.: 0.007 min
Response: 232732
Conc: 115.86 ng/ml m




#29 AR-1254-4

R.T.: 7.426 min
Delta R.T.: -0.014 min
Response: 362045
Conc: 141.14 ng/ml m



#29 AR-1254-4

R.T.: 6.339 min
Delta R.T.: -0.007 min
Response: 452204
Conc: 366.44 ng/ml




#30 AR-1254-5

R.T.: 7.868 min
Delta R.T.: 0.005 min
Response: 181663
Conc: 65.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
TRE-31702

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/08/2022
Supervised By :Ankita Jodhani 06/08/2022



#30 AR-1254-5

R.T.: 6.769 min
Delta R.T.: 0.004 min
Response: 172431
Conc: 99.55 ng/ml