

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060821\
 Data File : P0078551.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 08 Jun 2021 14:05
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 09 02:28:14 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.338	3.613	3724992	1571380	48.769	49.739
2)	SA Decachlor...	9.948	8.800	2109724	1361619	52.031	56.627
Target Compounds							
3)	L1 AR-1016-1	5.634	4.842	1115239	506535	476.284	484.214
4)	L1 AR-1016-2	5.656	4.861	1704918	745068	472.897	471.951
5)	L1 AR-1016-3	5.720	5.049	1057003	393994	483.321	492.017
6)	L1 AR-1016-4	5.824	5.101	849381	349695	484.121	495.177
7)	L1 AR-1016-5	6.136	5.324	839304	425649	483.797	488.011
31)	L7 AR-1260-1	7.296	6.403	1400337	758636	477.600	478.136
32)	L7 AR-1260-2	7.560	6.600	1633172	910300	496.464	488.747
33)	L7 AR-1260-3	7.920	6.751	1242641	849039	524.762	478.502
34)	L7 AR-1260-4	8.146	7.228	1378581	718605	536.002	490.393
35)	L7 AR-1260-5	8.460	7.476	2538658	1628841	499.218	466.815



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060821\
Data File : P0078551.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 08 Jun 2021 14:05
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 09 02:28:14 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 05:39:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.338 min
Delta R.T.: -0.001 min
Response: 3724992
Conc: 48.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.613 min
Delta R.T.: -0.002 min
Response: 1571380
Conc: 49.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.948 min
Delta R.T.: -0.002 min
Response: 2109724
Conc: 52.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.800 min
Delta R.T.: -0.003 min
Response: 1361619
Conc: 56.63 ng/ml



#3 AR-1016-1


R.T.: 5.634 min
Delta R.T.: -0.001 min
Response: 1115239
Conc: 476.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#3 AR-1016-1

R.T.: 4.842 min
Delta R.T.: -0.002 min
Response: 506535
Conc: 484.21 ng/ml




#4 AR-1016-2

R.T.: 5.656 min
Delta R.T.: 0.000 min
Response: 1704918
Conc: 472.90 ng/ml



#4 AR-1016-2


R.T.: 4.861 min
Delta R.T.: -0.002 min
Response: 745068
Conc: 471.95 ng/ml



#5 AR-1016-3


R.T.: 5.720 min
Delta R.T.: -0.001 min
Response: 1057003
Conc: 483.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#5 AR-1016-3

R.T.: 5.049 min
Delta R.T.: -0.001 min
Response: 393994
Conc: 492.02 ng/ml




#6 AR-1016-4

R.T.: 5.824 min
Delta R.T.: -0.002 min
Response: 849381
Conc: 484.12 ng/ml



#6 AR-1016-4


R.T.: 5.101 min
Delta R.T.: -0.003 min
Response: 349695
Conc: 495.18 ng/ml



#7 AR-1016-5


R.T.: 6.136 min
Delta R.T.: -0.001 min
Response: 839304
Conc: 483.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#7 AR-1016-5

R.T.: 5.324 min
Delta R.T.: -0.003 min
Response: 425649
Conc: 488.01 ng/ml




#31 AR-1260-1

R.T.: 7.296 min
Delta R.T.: -0.001 min
Response: 1400337
Conc: 477.60 ng/ml



#31 AR-1260-1


R.T.: 6.403 min
Delta R.T.: -0.003 min
Response: 758636
Conc: 478.14 ng/ml



#32 AR-1260-2


R.T.: 7.560 min
Delta R.T.: -0.002 min
Response: 1633172
Conc: 496.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#32 AR-1260-2

R.T.: 6.600 min
Delta R.T.: -0.004 min
Response: 910300
Conc: 488.75 ng/ml




#33 AR-1260-3

R.T.: 7.920 min
Delta R.T.: -0.002 min
Response: 1242641
Conc: 524.76 ng/ml



#33 AR-1260-3


R.T.: 6.751 min
Delta R.T.: -0.003 min
Response: 849039
Conc: 478.50 ng/ml



#34 AR-1260-4


R.T.: 8.146 min
Delta R.T.: -0.002 min
Response: 1378581
Conc: 536.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#34 AR-1260-4

R.T.: 7.228 min
Delta R.T.: -0.003 min
Response: 718605
Conc: 490.39 ng/ml



#35 AR-1260-5

R.T.: 8.460 min
Delta R.T.: 0.000 min
Response: 2538658
Conc: 499.22 ng/ml



#35 AR-1260-5

R.T.: 7.476 min
Delta R.T.: -0.003 min
Response: 1628841
Conc: 466.82 ng/ml