

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060921\
 Data File : P0078578.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 09 Jun 2021 12:57
 Operator : DD\AJ
 Sample : PB136944BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB136944BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 02:00:52 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.351	3.611	1584700	685626	20.747	21.702
2) SA Decachlor...	9.961	8.802	993858	599612	24.511	24.937

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060921\
Data File : P0078578.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 09 Jun 2021 12:57
Operator : DD\AJ
Sample : PB136944BL
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB136944BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 02:00:52 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 05:39:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.351 min
Delta R.T.: 0.012 min
Response: 1584700
Conc: 20.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB136944BL




#1 Tetrachloro-m-xylene

R.T.: 3.611 min
Delta R.T.: -0.004 min
Response: 685626
Conc: 21.70 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.961 min
Delta R.T.: 0.012 min
Response: 993858
Conc: 24.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.802 min
Delta R.T.: -0.001 min
Response: 599612
Conc: 24.94 ng/ml