

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060923\
 Data File : P0095582.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2023 16:54
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 06:29:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 10 06:28:38 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.405	3.622	403.2E6	172.8E6	95.580	93.370
2) SA Decachlor...	10.241	8.677	417.7E6	184.0E6	93.513	93.168
Target Compounds						
41) L9 AR-1268-1	8.699	7.562	338.3E6	151.7E6	950.267	958.649
42) L9 AR-1268-2	8.795	7.627	308.6E6	136.3E6	951.019	953.680
43) L9 AR-1268-3	9.030	7.836	270.7E6	130.7E6	946.510	944.690
44) L9 AR-1268-4	9.464	8.124	111.2E6	44813240	951.429	926.032
45) L9 AR-1268-5	9.892	8.419	900.5E6	387.8E6	988.599	969.732



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060923\
Data File : P0095582.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2023 16:54
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 06:29:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 10 06:28:38 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.001 min
Response: 403239571
Conc: 95.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: 0.000 min
Response: 172805609
Conc: 93.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: 0.000 min
Response: 417651352
Conc: 93.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: 0.000 min
Response: 184038973
Conc: 93.17 ng/ml



#41 AR-1268-1


R.T.: 8.699 min
Delta R.T.: 0.000 min
Response: 338346823
Conc: 950.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.562 min
Delta R.T.: 0.000 min
Response: 151696426
Conc: 958.65 ng/ml




#42 AR-1268-2

R.T.: 8.795 min
Delta R.T.: 0.000 min
Response: 308647061
Conc: 951.02 ng/ml



#42 AR-1268-2


R.T.: 7.627 min
Delta R.T.: 0.000 min
Response: 136335934
Conc: 953.68 ng/ml



#43 AR-1268-3


R.T.: 9.030 min
Delta R.T.: 0.001 min
Response: 270728704
Conc: 946.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.836 min
Delta R.T.: 0.000 min
Response: 130698461
Conc: 944.69 ng/ml




#44 AR-1268-4

R.T.: 9.464 min
Delta R.T.: 0.001 min
Response: 111212238
Conc: 951.43 ng/ml



#44 AR-1268-4


R.T.: 8.124 min
Delta R.T.: 0.000 min
Response: 44813240
Conc: 926.03 ng/ml



#45 AR-1268-5

R.T.: 9.892 min
Delta R.T.: 0.001 min
Response: 900539326
Conc: 988.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.419 min
Delta R.T.: 0.001 min
Response: 387775831
Conc: 969.73 ng/ml