

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060923\
 Data File : P0095566.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 Jun 2023 12:23
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 10 01:37:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 10 01:37:39 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.406	3.622	375.4E6	159.8E6	98.766	99.318
2) SA Decachlor...	10.238	8.677	233.3E6	99986470	95.586	95.273
Target Compounds						
16) L4 AR-1242-1	5.587	4.716	78051004	29457151	968.559	968.854
17) L4 AR-1242-2	5.610	4.735	112.9E6	41907511	970.917	974.469
18) L4 AR-1242-3	5.672	4.912	74145545	22848845	963.480	964.332
19) L4 AR-1242-4	5.772	4.997	58890807	26911600	970.887	964.751
20) L4 AR-1242-5	6.516	5.524	55359103	27488465	956.794	969.255



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0060923\
Data File : P0095566.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 Jun 2023 12:23
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 10 01:37:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 10 01:37:39 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.406 min
Delta R.T.: 0.000 min
Response: 375400286
Conc: 98.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: 0.000 min
Response: 159766623
Conc: 99.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.238 min
Delta R.T.: 0.000 min
Response: 233285205
Conc: 95.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: 0.000 min
Response: 99986470
Conc: 95.27 ng/ml



#16 AR-1242-1


R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 78051004
Conc: 968.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.716 min
Delta R.T.: 0.000 min
Response: 29457151
Conc: 968.85 ng/ml




#17 AR-1242-2

R.T.: 5.610 min
Delta R.T.: 0.000 min
Response: 112940514
Conc: 970.92 ng/ml



#17 AR-1242-2


R.T.: 4.735 min
Delta R.T.: 0.000 min
Response: 41907511
Conc: 974.47 ng/ml



#18 AR-1242-3


R.T.: 5.672 min
Delta R.T.: 0.000 min
Response: 74145545
Conc: 963.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 4.912 min
Delta R.T.: 0.000 min
Response: 22848845
Conc: 964.33 ng/ml




#19 AR-1242-4

R.T.: 5.772 min
Delta R.T.: 0.000 min
Response: 58890807
Conc: 970.89 ng/ml



#19 AR-1242-4


R.T.: 4.997 min
Delta R.T.: 0.000 min
Response: 26911600
Conc: 964.75 ng/ml



#20 AR-1242-5

R.T.: 6.516 min
Delta R.T.: 0.000 min
Response: 55359103
Conc: 956.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.524 min
Delta R.T.: 0.000 min
Response: 27488465
Conc: 969.26 ng/ml