

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061019\
 Data File : P0057033.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Jun 2019 13:08
 Operator : SM/SJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Manual Integrations
 APPROVED

Ankita
 6/11/2019 2:45:20 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 05:47:31 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0052319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 24 00:11:05 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.245	3.563	2333215	1962986	60.946	59.523
2)	SA Decachlor...	9.811	8.494	2378402	1671048	48.456	46.865
Target Compounds							
3)	L1 AR-1016-1	5.405	4.630	974633	696060	541.878m	553.093m
4)	L1 AR-1016-2	5.427	4.649	1289584	921153	513.559m	511.290m
5)	L1 AR-1016-3	5.489	4.824	820551	534020	510.516	541.632
6)	L1 AR-1016-4	5.587	4.864	676157	434343	515.442	524.669
7)	L1 AR-1016-5	5.877	5.074	698675	558398	505.697	520.543m
31)	L7 AR-1260-1	6.993	6.095	1235120	976399	481.307	488.106
32)	L7 AR-1260-2	7.249	6.283	1489576	1222241	492.558	495.846
33)	L7 AR-1260-3	7.605	6.435	1202517	1101943	499.166	491.231
34)	L7 AR-1260-4	7.830	6.901	1252657	920985	478.257	494.882
35)	L7 AR-1260-5	8.138	7.143	2374712	2250982	496.585	521.987

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061019\
Data File : PO057033.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Jun 2019 13:08
Operator : SM/SJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 05:47:31 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO052319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 24 00:11:05 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.245 min
Delta R.T.: -0.011 min
Response: 2333215
Conc: 60.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM




#1 Tetrachloro-m-xylene

R.T.: 3.563 min
Delta R.T.: -0.010 min
Response: 1962986
Conc: 59.52 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.811 min
Delta R.T.: -0.025 min
Response: 2378402
Conc: 48.46 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.494 min
Delta R.T.: -0.019 min
Response: 1671048
Conc: 46.86 ng/ml




#3 AR-1016-1

R.T.: 5.405 min
Delta R.T.: -0.013 min
Response: 974633
Conc: 541.88 ng/ml m

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM




#3 AR-1016-1

R.T.: 4.630 min
Delta R.T.: -0.013 min
Response: 696060
Conc: 553.09 ng/ml m




#4 AR-1016-2

R.T.: 5.427 min
Delta R.T.: -0.012 min
Response: 1289584
Conc: 513.56 ng/ml m



#4 AR-1016-2

R.T.: 4.649 min
Delta R.T.: -0.013 min
Response: 921153
Conc: 511.29 ng/ml m




#5 AR-1016-3

R.T.: 5.489 min
Delta R.T.: -0.012 min
Response: 820551
Conc: 510.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM




#5 AR-1016-3

R.T.: 4.824 min
Delta R.T.: -0.012 min
Response: 534020
Conc: 541.63 ng/ml




#6 AR-1016-4

R.T.: 5.587 min
Delta R.T.: -0.012 min
Response: 676157
Conc: 515.44 ng/ml



#6 AR-1016-4

R.T.: 4.864 min
Delta R.T.: -0.013 min
Response: 434343
Conc: 524.67 ng/ml




#7 AR-1016-5

R.T.: 5.877 min
Delta R.T.: -0.013 min
Response: 698675
Conc: 505.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM




#7 AR-1016-5

R.T.: 5.074 min
Delta R.T.: -0.013 min
Response: 558398
Conc: 520.54 ng/ml m




#31 AR-1260-1

R.T.: 6.993 min
Delta R.T.: -0.015 min
Response: 1235120
Conc: 481.31 ng/ml



#31 AR-1260-1

R.T.: 6.095 min
Delta R.T.: -0.015 min
Response: 976399
Conc: 488.11 ng/ml




#32 AR-1260-2

R.T.: 7.249 min
Delta R.T.: -0.015 min
Response: 1489576
Conc: 492.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM




#32 AR-1260-2

R.T.: 6.283 min
Delta R.T.: -0.015 min
Response: 1222241
Conc: 495.85 ng/ml




#33 AR-1260-3

R.T.: 7.605 min
Delta R.T.: -0.015 min
Response: 1202517
Conc: 499.17 ng/ml



#33 AR-1260-3

R.T.: 6.435 min
Delta R.T.: -0.015 min
Response: 1101943
Conc: 491.23 ng/ml




#34 AR-1260-4

R.T.: 7.830 min
Delta R.T.: -0.016 min
Response: 1252657
Conc: 478.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500


Manual Integrations
APPROVED

Ankita
6/11/2019 2:45:20 PM




#34 AR-1260-4

R.T.: 6.901 min
Delta R.T.: -0.016 min
Response: 920985
Conc: 494.88 ng/ml



#35 AR-1260-5

R.T.: 8.138 min
Delta R.T.: -0.016 min
Response: 2374712
Conc: 496.58 ng/ml



#35 AR-1260-5

R.T.: 7.143 min
Delta R.T.: -0.015 min
Response: 2250982
Conc: 521.99 ng/ml