

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061021\
 Data File : P0078635.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 10 Jun 2021 22:04
 Operator : DD\AJ
 Sample : M2641-02
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HIB135S-END-1-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 11 05:19:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.338	3.613	803595	462833	10.521	14.650 #
2) SA Decachlor...	9.953	8.805	635706	405982	15.678	16.884

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061021\
Data File : P0078635.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 10 Jun 2021 22:04
Operator : DD\AJ
Sample : M2641-02
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HIB135S-END-1-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 11 05:19:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 05:39:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.338 min
Delta R.T.: 0.000 min
Response: 803595
Conc: 10.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
HIB135S-END-1-2




#1 Tetrachloro-m-xylene

R.T.: 3.613 min
Delta R.T.: -0.002 min
Response: 462833
Conc: 14.65 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.953 min
Delta R.T.: 0.003 min
Response: 635706
Conc: 15.68 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.805 min
Delta R.T.: 0.002 min
Response: 405982
Conc: 16.88 ng/ml