

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061121\
 Data File : P0078705.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jun 2021 6:30
 Operator : DD\AJ
 Sample : M2656-01
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 FM628

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 07:06:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jun 08 05:39:51 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.334	3.611	1192360	525675	15.611	16.639
2) SA Decachlor...	9.943	8.793	591415	341992	14.586	14.223

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061121\
Data File : P0078705.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 Jun 2021 6:30
Operator : DD\AJ
Sample : M2656-01
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
FM628

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 07:06:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060721.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jun 08 05:39:51 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.334 min
Delta R.T.: -0.005 min
Response: 1192360
Conc: 15.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
FM628




#1 Tetrachloro-m-xylene

R.T.: 3.611 min
Delta R.T.: -0.004 min
Response: 525675
Conc: 16.64 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.943 min
Delta R.T.: -0.007 min
Response: 591415
Conc: 14.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.793 min
Delta R.T.: -0.009 min
Response: 341992
Conc: 14.22 ng/ml