

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061124\
 Data File : P0104277.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 11 Jun 2024 15:29
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 04:23:33 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 12 04:06:39 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.401	3.546	835.1E6	224.0E6	97.940	98.920
2) SA Decachlor...	10.104	8.460	565.8E6	308.8E6	96.313	96.447
Target Compounds						
21) L5 AR-1248-1	5.563	4.606	166.5E6	42478773	961.169	955.877
22) L5 AR-1248-2	5.836	4.838	234.9E6	52307848	953.442	950.605
23) L5 AR-1248-3	6.040	4.878	253.8E6	58895345	957.619	954.763
24) L5 AR-1248-4	6.442	5.047	281.2E6	73103612	964.611	960.081
25) L5 AR-1248-5	6.481	5.432	272.4E6	79993849	965.469	967.079



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061124\
Data File : PO104277.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 11 Jun 2024 15:29
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 04:23:33 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061124.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 12 04:06:39 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.401 min
Delta R.T.: 0.000 min
Response: 835120621
Conc: 97.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: 0.000 min
Response: 223970070
Conc: 98.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.104 min
Delta R.T.: 0.000 min
Response: 565763148
Conc: 96.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.460 min
Delta R.T.: 0.000 min
Response: 308826842
Conc: 96.45 ng/ml



#21 AR-1248-1


R.T.: 5.563 min
Delta R.T.: 0.000 min
Response: 166496413
Conc: 961.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.606 min
Delta R.T.: 0.000 min
Response: 42478773
Conc: 955.88 ng/ml




#22 AR-1248-2

R.T.: 5.836 min
Delta R.T.: 0.000 min
Response: 234851677
Conc: 953.44 ng/ml



#22 AR-1248-2


R.T.: 4.838 min
Delta R.T.: 0.000 min
Response: 52307848
Conc: 950.61 ng/ml



#23 AR-1248-3


R.T.: 6.040 min
Delta R.T.: 0.000 min
Response: 253761897
Conc: 957.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 58895345
Conc: 954.76 ng/ml




#24 AR-1248-4

R.T.: 6.442 min
Delta R.T.: 0.000 min
Response: 281204504
Conc: 964.61 ng/ml



#24 AR-1248-4


R.T.: 5.047 min
Delta R.T.: 0.000 min
Response: 73103612
Conc: 960.08 ng/ml



#25 AR-1248-5

R.T.: 6.481 min
Delta R.T.: 0.000 min
Response: 272360130
Conc: 965.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.432 min
Delta R.T.: 0.000 min
Response: 79993849
Conc: 967.08 ng/ml