

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061220\
 Data File : P0068918.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jun 2020 12:59
 Operator : DD\AJ
 Sample : P0061220ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 17:32:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 12 17:24:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	27.285	26.130 E3	4.2	96	0.00
2 SA Decachlorobiphenyl	44.906	45.622 E3	-1.6	100	0.00
3 L1 AR-1016-1	1.289	1.232 E3	4.4	98	0.00
4 L1 AR-1016-2	1.817	1.742 E3	4.1	97	0.00
5 L1 AR-1016-3	1.134	1.087 E3	4.1	98	0.00
6 L1 AR-1016-4	922.983	907.135	1.7	98	0.00
7 L1 AR-1016-5	949.496	922.143	2.9	99	0.00
31 L7 AR-1260-1	1.807	1.756 E3	2.8	99	0.00
32 L7 AR-1260-2	2.238	2.177 E3	2.7	99	0.00
33 L7 AR-1260-3	1.738	1.687 E3	2.9	100	0.00
34 L7 AR-1260-4	2.039	1.955 E3	4.1	99	0.00
35 L7 AR-1260-5	4.455	4.260 E3	4.4	100	0.00

Signal #2

1 SA Tetrachloro-m-xylene	35.527	34.138 E3	3.9	96	0.00
2 SA Decachlorobiphenyl	51.172	51.869 E3	-1.4	101	0.00
3 L1 AR-1016-1	1.392	1.368 E3	1.7	98	0.00
4 L1 AR-1016-2	1.958	1.890 E3	3.5	96	0.00
5 L1 AR-1016-3	1.031	0.999 E3	3.1	97	0.00
6 L1 AR-1016-4	858.632	852.429	0.7	97	0.00
7 L1 AR-1016-5	1.053	1.064 E3	-1.0	99	0.00
31 L7 AR-1260-1	2.101	2.070 E3	1.5	99	0.00
32 L7 AR-1260-2	2.722	2.642 E3	2.9	99	0.00
33 L7 AR-1260-3	2.430	2.421 E3	0.4	99	0.00
34 L7 AR-1260-4	2.134	2.170 E3	-1.7	99	0.00
35 L7 AR-1260-5	5.775	5.618 E3	2.7	99	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	327.719	0.000	100.0#	0#	-4.66#
9 L2 AR-1221-2	242.385	0.000	100.0#	0#	-4.76#
10 L2 AR-1221-3	823.712	0.000	100.0#	0#	-4.85#
11 L3 AR-1232-1	700.150	0.000	100.0#	0#	-4.85#
12 L3 AR-1232-2	395.419	0.000	100.0#	0#	-5.42#
13 L3 AR-1232-3	805.526	0.000	100.0#	0#	-5.74#
14 L3 AR-1232-4	406.969	0.000	100.0#	0#	-5.91#
15 L3 AR-1232-5	277.204	0.000	100.0#	0#	-6.01#
16 L4 AR-1242-1	1.010	0.000 E3	100.0#	0#	-5.72#
17 L4 AR-1242-2	1.410	0.000 E3	100.0#	0#	-5.74#
18 L4 AR-1242-3	887.916	0.000	100.0#	0#	-5.80#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061220\
 Data File : P0068918.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jun 2020 12:59
 Operator : DD\AJ
 Sample : P0061220ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 17:32:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 12 17:24:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	738.527	0.000	100.0#	0#	-5.91#
20 L4 AR-1242-5	855.247	0.000	100.0#	0#	-6.68#
21 L5 AR-1248-1	786.702	0.000	100.0#	0#	-5.72#
22 L5 AR-1248-2	1.007	0.000 E3	100.0#	0#	-6.01#
23 L5 AR-1248-3	1.233	0.000 E3	100.0#	0#	-6.22#
24 L5 AR-1248-4	1.472	0.000 E3	100.0#	0#	-6.65#
25 L5 AR-1248-5	1.412	0.000 E3	100.0#	0#	-6.69#
26 L6 AR-1254-1	1.457	0.000 E3	100.0#	0#	-6.61#
27 L6 AR-1254-2	2.329	0.000 E3	100.0#	0#	-6.84#
28 L6 AR-1254-3	2.561	0.000 E3	100.0#	0#	-7.22#
29 L6 AR-1254-4	2.030	0.000 E3	100.0#	0#	-7.51#
30 L6 AR-1254-5	2.170	0.000 E3	100.0#	0#	-7.94#
36 L8 AR-1262-1	2.539	0.000 E3	100.0#	0#	-8.01#
37 L8 AR-1262-2	4.932	0.000 E3	100.0#	0#	-8.55#
38 L8 AR-1262-3	2.216	0.000 E3	100.0#	0#	-8.82#
39 L8 AR-1262-4	1.409	0.000 E3	100.0#	0#	-8.90#
40 L8 AR-1262-5	1.881	0.000 E3	100.0#	0#	-9.47#
41 L9 AR-1268-1	6.049	0.000 E3	100.0#	0#	-8.82#
42 L9 AR-1268-2	5.809	0.000 E3	100.0#	0#	-8.91#
43 L9 AR-1268-3	4.856	0.000 E3	100.0#	0#	-9.09#
44 L9 AR-1268-4	2.178	0.000 E3	100.0#	0#	-9.47#
45 L9 AR-1268-5	17.314	0.000 E3	100.0#	0#	-9.81#

Signal #2

8 L2 AR-1221-1	412.506	0.000	100.0#	0#	-3.82#
9 L2 AR-1221-2	307.121	0.000	100.0#	0#	-3.92#
10 L2 AR-1221-3	972.324	0.000	100.0#	0#	-4.00#
11 L3 AR-1232-1	801.495	0.000	100.0#	0#	-4.00#
12 L3 AR-1232-2	846.660	0.000	100.0#	0#	-4.82#
13 L3 AR-1232-3	436.437	0.000	100.0#	0#	-5.00#
14 L3 AR-1232-4	360.354	0.000	100.0#	0#	-5.10#
15 L3 AR-1232-5	418.384	0.000	100.0#	0#	-5.28#
16 L4 AR-1242-1	1.076	0.000 E3	100.0#	0#	-4.80#
17 L4 AR-1242-2	1.526	0.000 E3	100.0#	0#	-4.82#
18 L4 AR-1242-3	798.739	0.000	100.0#	0#	-5.00#
19 L4 AR-1242-4	732.168	0.000	100.0#	0#	-5.10#
20 L4 AR-1242-5	1.045	0.000 E3	100.0#	0#	-5.66#
21 L5 AR-1248-1	821.072	0.000	100.0#	0#	-4.80#
22 L5 AR-1248-2	1.155	0.000 E3	100.0#	0#	-5.06#
23 L5 AR-1248-3	1.076	0.000 E3	100.0#	0#	-5.10#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061220\
 Data File : P0068918.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 Jun 2020 12:59
 Operator : DD\AJ
 Sample : P0061220ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 17:32:09 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061220.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 12 17:24:41 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.388	0.000 E3	100.0#	0#	-5.28#
25 L5 AR-1248-5	1.429	0.000 E3	100.0#	0#	-5.70#
26 L6 AR-1254-1	2.264	0.000 E3	100.0#	0#	-5.66#
27 L6 AR-1254-2	2.004	0.000 E3	100.0#	0#	-5.82#
28 L6 AR-1254-3	3.312	0.000 E3	100.0#	0#	-6.23#
29 L6 AR-1254-4	2.228	0.000 E3	100.0#	0#	-6.47#
30 L6 AR-1254-5	3.217	0.000 E3	100.0#	0#	-6.89#
36 L8 AR-1262-1	1.712	0.000 E3	100.0#	0#	-6.89#
37 L8 AR-1262-2	6.351	0.000 E3	100.0#	0#	-7.44#
38 L8 AR-1262-3	2.485	0.000 E3	100.0#	0#	-7.73#
39 L8 AR-1262-4	4.801	0.000 E3	100.0#	0#	-7.79#
40 L8 AR-1262-5	2.343	0.000 E3	100.0#	0#	-8.29#
41 L9 AR-1268-1	7.602	0.000 E3	100.0#	0#	-7.73#
42 L9 AR-1268-2	7.128	0.000 E3	100.0#	0#	-7.79#
43 L9 AR-1268-3	5.839	0.000 E3	100.0#	0#	-8.00#
44 L9 AR-1268-4	2.695	0.000 E3	100.0#	0#	-8.29#
45 L9 AR-1268-5	19.977	0.000 E3	100.0#	0#	-8.56#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0