

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061223\
 Data File : P0095609.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 12 Jun 2023 16:26
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 12 22:42:14 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 10 06:51:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.407	3.625	185.1E6	79636922	43.193	43.806
2) SA Decachlor...	10.247	8.682	129.5E6	59782871	51.962	53.679
Target Compounds						
21) L5 AR-1248-1	5.589	4.720	33201006	13632421	492.661	530.535
22) L5 AR-1248-2	5.866	4.959	51000800	22092777	471.155	504.383
23) L5 AR-1248-3	6.072	5.001	52998566	22885126	479.417	507.058
24) L5 AR-1248-4	6.479	5.174	53769173	25751634	489.442	507.236
25) L5 AR-1248-5	6.519	5.569	54467927	20833116	484.478	540.976



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061223\
Data File : P0095609.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 12 Jun 2023 16:26
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 12 22:42:14 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 10 06:51:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.407 min
Delta R.T.: 0.004 min
Response: 185131731
Conc: 43.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.625 min
Delta R.T.: 0.006 min
Response: 79636922
Conc: 43.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.247 min
Delta R.T.: 0.012 min
Response: 129504103
Conc: 51.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.682 min
Delta R.T.: 0.007 min
Response: 59782871
Conc: 53.68 ng/ml



#21 AR-1248-1


R.T.: 5.589 min
Delta R.T.: 0.004 min
Response: 33201006
Conc: 492.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.720 min
Delta R.T.: 0.006 min
Response: 13632421
Conc: 530.53 ng/ml




#22 AR-1248-2

R.T.: 5.866 min
Delta R.T.: 0.004 min
Response: 51000800
Conc: 471.15 ng/ml



#22 AR-1248-2


R.T.: 4.959 min
Delta R.T.: 0.006 min
Response: 22092777
Conc: 504.38 ng/ml



#23 AR-1248-3


R.T.: 6.072 min
Delta R.T.: 0.004 min
Response: 52998566
Conc: 479.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.001 min
Delta R.T.: 0.006 min
Response: 22885126
Conc: 507.06 ng/ml




#24 AR-1248-4

R.T.: 6.479 min
Delta R.T.: 0.005 min
Response: 53769173
Conc: 489.44 ng/ml



#24 AR-1248-4


R.T.: 5.174 min
Delta R.T.: 0.006 min
Response: 25751634
Conc: 507.24 ng/ml



#25 AR-1248-5

R.T.: 6.519 min
Delta R.T.: 0.005 min
Response: 54467927
Conc: 484.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.569 min
Delta R.T.: 0.006 min
Response: 20833116
Conc: 540.98 ng/ml