

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061321\
 Data File : P0078738.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jun 2021 17:02
 Operator : DD\AJ
 Sample : P0061321ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 05:36:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	84.667	86.271 E3	-1.9	101	0.00
2 SA Decachlorobiphenyl	58.684	59.408 E3	-1.2	102	0.00
3 L1 AR-1016-1	2.553	2.557 E3	-0.2	103	0.00
4 L1 AR-1016-2	3.897	3.986 E3	-2.3	104	0.00
5 L1 AR-1016-3	2.356	2.566 E3	-8.9	109	0.00
6 L1 AR-1016-4	1.889	2.060 E3	-9.1	111	0.00
7 L1 AR-1016-5	1.792	1.943 E3	-8.4	103	0.00
31 L7 AR-1260-1	3.144	3.280 E3	-4.3	106	0.00
32 L7 AR-1260-2	3.697	3.773 E3	-2.1	104	0.00
33 L7 AR-1260-3	2.836	2.895 E3	-2.1	104	0.00
34 L7 AR-1260-4	3.159	3.272 E3	-3.6	104	0.00
35 L7 AR-1260-5	6.103	6.084 E3	0.3	102	0.00

Signal #2

1 SA Tetrachloro-m-xylene	34.618	33.107 E3	4.4	96	0.00
2 SA Decachlorobiphenyl	31.092	30.954 E3	0.4	103	0.00
3 L1 AR-1016-1	1.131	1.094 E3	3.3	102	0.00
4 L1 AR-1016-2	1.657	1.645 E3	0.7	102	0.00
5 L1 AR-1016-3	858.642	864.234	-0.7	103	0.00
6 L1 AR-1016-4	755.150	755.811	-0.1	102	0.00
7 L1 AR-1016-5	911.070	925.131	-1.5	104	0.00
31 L7 AR-1260-1	1.680	1.656 E3	1.4	103	0.00
32 L7 AR-1260-2	2.015	1.995 E3	1.0	102	0.00
33 L7 AR-1260-3	1.951	1.862 E3	4.6	102	0.00
34 L7 AR-1260-4	1.598	1.579 E3	1.2	102	0.00
35 L7 AR-1260-5	3.679	3.651 E3	0.8	102	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	899.884	0.000	100.0#	0#	-4.58#
9 L2 AR-1221-2	663.498	0.000	100.0#	0#	-4.68#
10 L2 AR-1221-3	2.131	0.000 E3	100.0#	0#	-4.76#
11 L3 AR-1232-1	1.665	0.000 E3	100.0#	0#	-4.76#
12 L3 AR-1232-2	819.510	0.000	100.0#	0#	-5.34#
13 L3 AR-1232-3	1.595	0.000 E3	100.0#	0#	-5.65#
14 L3 AR-1232-4	749.547	0.000	100.0#	0#	-5.82#
15 L3 AR-1232-5	522.568	0.000	100.0#	0#	-5.92#
16 L4 AR-1242-1	1.970	0.000 E3	100.0#	0#	-5.63#
17 L4 AR-1242-2	3.019	0.000 E3	100.0#	0#	-5.65#
18 L4 AR-1242-3	1.826	0.000 E3	100.0#	0#	-5.71#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061321\
 Data File : P0078738.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jun 2021 17:02
 Operator : DD\AJ
 Sample : P0061321ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 05:36:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound			AvgRF	CCRF	%Dev	Area%	Dev(Min)

19	L4	AR-1242-4	1.474	0.000 E3	100.0#	0#	-5.82#
20	L4	AR-1242-5	1.504	0.000 E3	100.0#	0#	-6.59#
21	L5	AR-1248-1	1.536	0.000 E3	100.0#	0#	-5.63#
22	L5	AR-1248-2	2.087	0.000 E3	100.0#	0#	-5.92#
23	L5	AR-1248-3	2.455	0.000 E3	100.0#	0#	-6.13#
24	L5	AR-1248-4	2.619	0.000 E3	100.0#	0#	-6.55#
25	L5	AR-1248-5	2.527	0.000 E3	100.0#	0#	-6.59#
26	L6	AR-1254-1	2.741	0.000 E3	100.0#	0#	-6.52#
27	L6	AR-1254-2	4.230	0.000 E3	100.0#	0#	-6.75#
28	L6	AR-1254-3	4.354	0.000 E3	100.0#	0#	-7.12#
29	L6	AR-1254-4	3.078	0.000 E3	100.0#	0#	-7.42#
30	L6	AR-1254-5	3.447	0.000 E3	100.0#	0#	-7.84#
36	L8	AR-1262-1	4.142	0.000 E3	100.0#	0#	-7.91#
37	L8	AR-1262-2	6.910	0.000 E3	100.0#	0#	-8.45#
38	L8	AR-1262-3	4.732	0.000 E3	100.0#	0#	-8.72#
39	L8	AR-1262-4	3.618	0.000 E3	100.0#	0#	-8.80#
40	L8	AR-1262-5	2.443	0.000 E3	100.0#	0#	-9.35#
41	L9	AR-1268-1	7.864	0.000 E3	100.0#	0#	-8.72#
42	L9	AR-1268-2	7.219	0.000 E3	100.0#	0#	-8.80#
43	L9	AR-1268-3	6.185	0.000 E3	100.0#	0#	-8.98#
44	L9	AR-1268-4	2.574	0.000 E3	100.0#	0#	-9.35#
45	L9	AR-1268-5	19.702	0.000 E3	100.0#	0#	-9.68#

Signal #2

8	L2	AR-1221-1	377.539	0.000	100.0#	0#	-3.86#
9	L2	AR-1221-2	286.693	0.000	100.0#	0#	-3.96#
10	L2	AR-1221-3	904.563	0.000	100.0#	0#	-4.04#
11	L3	AR-1232-1	718.198	0.000	100.0#	0#	-4.04#
12	L3	AR-1232-2	704.266	0.000	100.0#	0#	-4.85#
13	L3	AR-1232-3	349.486	0.000	100.0#	0#	-5.04#
14	L3	AR-1232-4	325.787	0.000	100.0#	0#	-5.14#
15	L3	AR-1232-5	350.954	0.000	100.0#	0#	-5.32#
16	L4	AR-1242-1	875.114	0.000	100.0#	0#	-4.83#
17	L4	AR-1242-2	1.293	0.000 E3	100.0#	0#	-4.85#
18	L4	AR-1242-3	678.828	0.000	100.0#	0#	-5.04#
19	L4	AR-1242-4	698.052	0.000	100.0#	0#	-5.14#
20	L4	AR-1242-5	894.623	0.000	100.0#	0#	-5.69#
21	L5	AR-1248-1	677.108	0.000	100.0#	0#	-4.83#
22	L5	AR-1248-2	1.009	0.000 E3	100.0#	0#	-5.09#
23	L5	AR-1248-3	998.074	0.000	100.0#	0#	-5.13#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061321\
 Data File : P0078738.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 13 Jun 2021 17:02
 Operator : DD\AJ
 Sample : P0061321ICV500
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 05:36:16 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	1.183	0.000 E3	100.0#	0#	-5.32#
25 L5 AR-1248-5	1.102	0.000 E3	100.0#	0#	-5.73#
26 L6 AR-1254-1	1.767	0.000 E3	100.0#	0#	-5.69#
27 L6 AR-1254-2	1.620	0.000 E3	100.0#	0#	-5.85#
28 L6 AR-1254-3	2.431	0.000 E3	100.0#	0#	-6.26#
29 L6 AR-1254-4	1.497	0.000 E3	100.0#	0#	-6.50#
30 L6 AR-1254-5	2.296	0.000 E3	100.0#	0#	-6.92#
36 L8 AR-1262-1	1.190	0.000 E3	100.0#	0#	-6.92#
37 L8 AR-1262-2	3.951	0.000 E3	100.0#	0#	-7.47#
38 L8 AR-1262-3	1.653	0.000 E3	100.0#	0#	-7.75#
39 L8 AR-1262-4	3.045	0.000 E3	100.0#	0#	-7.81#
40 L8 AR-1262-5	1.425	0.000 E3	100.0#	0#	-8.31#
41 L9 AR-1268-1	4.399	0.000 E3	100.0#	0#	-7.75#
42 L9 AR-1268-2	3.973	0.000 E3	100.0#	0#	-7.81#
43 L9 AR-1268-3	3.440	0.000 E3	100.0#	0#	-8.01#
44 L9 AR-1268-4	1.469	0.000 E3	100.0#	0#	-8.31#
45 L9 AR-1268-5	10.356	0.000 E3	100.0#	0#	-8.57#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0