

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061322\
 Data File : P0087102.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2022 10:00
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 11:00:45 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.492	3.639	121.5E6	56790543	52.007	50.439
2) SA Decachlor...	10.352	8.649	92904221	42714449	50.482	54.327
Target Compounds						
16) L4 AR-1242-1	5.676	4.729	32013732	12310860	498.081	465.125
17) L4 AR-1242-2	5.699	4.748	45503844	16844095	497.998	467.622
18) L4 AR-1242-3	5.762	4.924	28656472	9133138	488.218	457.049
19) L4 AR-1242-4	5.862	5.008	23147337	8816152	496.856	451.044
20) L4 AR-1242-5	6.604	5.530	23623546	10900189	497.878	466.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061322\
Data File : P0087102.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2022 10:00
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 11:00:45 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.492 min
Delta R.T.: 0.000 min
Response: 121540891
Conc: 52.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 56790543
Conc: 50.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.352 min
Delta R.T.: -0.003 min
Response: 92904221
Conc: 50.48 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.649 min
Delta R.T.: -0.008 min
Response: 42714449
Conc: 54.33 ng/ml



#16 AR-1242-1


R.T.: 5.676 min
Delta R.T.: 0.000 min
Response: 32013732
Conc: 498.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.729 min
Delta R.T.: -0.003 min
Response: 12310860
Conc: 465.13 ng/ml




#17 AR-1242-2

R.T.: 5.699 min
Delta R.T.: 0.000 min
Response: 45503844
Conc: 498.00 ng/ml



#17 AR-1242-2


R.T.: 4.748 min
Delta R.T.: -0.003 min
Response: 16844095
Conc: 467.62 ng/ml



#18 AR-1242-3


R.T.: 5.762 min
Delta R.T.: 0.000 min
Response: 28656472
Conc: 488.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.924 min
Delta R.T.: -0.003 min
Response: 9133138
Conc: 457.05 ng/ml




#19 AR-1242-4

R.T.: 5.862 min
Delta R.T.: 0.000 min
Response: 23147337
Conc: 496.86 ng/ml



#19 AR-1242-4


R.T.: 5.008 min
Delta R.T.: -0.004 min
Response: 8816152
Conc: 451.04 ng/ml



#20 AR-1242-5

R.T.: 6.604 min
Delta R.T.: 0.000 min
Response: 23623546
Conc: 497.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.530 min
Delta R.T.: -0.004 min
Response: 10900189
Conc: 466.91 ng/ml