

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061322\
 Data File : P0087115.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2022 14:56
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 15:12:00 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.639	114.4E6	54596658	48.949	48.490
2) SA Decachlor...	10.350	8.651	88041492	41598337	47.840	52.907
Target Compounds						
3) L1 AR-1016-1	5.676	4.729	39204096	15479243	468.372	459.715
4) L1 AR-1016-2	5.698	4.748	55109794	21018955	468.671	451.859
5) L1 AR-1016-3	5.760	4.924	34722807	11479845	461.293	448.111
6) L1 AR-1016-4	5.861	4.967	27946743	9097460	466.412	446.249
7) L1 AR-1016-5	6.158	5.180	27809398	11253723	473.170	440.353
31) L7 AR-1260-1	7.294	6.213	46915860	21183572	465.558	452.245
32) L7 AR-1260-2	7.552	6.401	55103290	25453860	467.556	454.421
33) L7 AR-1260-3	7.914	6.554	39362805	23770176	464.330	446.701
34) L7 AR-1260-4	8.143	7.026	46316651	18887483	454.778	468.656
35) L7 AR-1260-5	8.471	7.266	85465205	44845563	470.243	478.070



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061322\
Data File : P0087115.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2022 14:56
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 15:12:00 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.002 min
Response: 114395331
Conc: 48.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 54596658
Conc: 48.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.350 min
Delta R.T.: -0.005 min
Response: 88041492
Conc: 47.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.651 min
Delta R.T.: -0.006 min
Response: 41598337
Conc: 52.91 ng/ml



#3 AR-1016-1


R.T.: 5.676 min
Delta R.T.: -0.002 min
Response: 39204096
Conc: 468.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.729 min
Delta R.T.: -0.003 min
Response: 15479243
Conc: 459.71 ng/ml




#4 AR-1016-2

R.T.: 5.698 min
Delta R.T.: -0.002 min
Response: 55109794
Conc: 468.67 ng/ml



#4 AR-1016-2


R.T.: 4.748 min
Delta R.T.: -0.003 min
Response: 21018955
Conc: 451.86 ng/ml



#5 AR-1016-3


R.T.: 5.760 min
Delta R.T.: -0.003 min
Response: 34722807
Conc: 461.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.924 min
Delta R.T.: -0.003 min
Response: 11479845
Conc: 448.11 ng/ml




#6 AR-1016-4

R.T.: 5.861 min
Delta R.T.: -0.002 min
Response: 27946743
Conc: 466.41 ng/ml



#6 AR-1016-4


R.T.: 4.967 min
Delta R.T.: -0.003 min
Response: 9097460
Conc: 446.25 ng/ml



#7 AR-1016-5


R.T.: 6.158 min
Delta R.T.: -0.003 min
Response: 27809398
Conc: 473.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.180 min
Delta R.T.: -0.003 min
Response: 11253723
Conc: 440.35 ng/ml




#31 AR-1260-1

R.T.: 7.294 min
Delta R.T.: -0.003 min
Response: 46915860
Conc: 465.56 ng/ml



#31 AR-1260-1


R.T.: 6.213 min
Delta R.T.: -0.004 min
Response: 21183572
Conc: 452.25 ng/ml



#32 AR-1260-2


R.T.: 7.552 min
Delta R.T.: -0.003 min
Response: 55103290
Conc: 467.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.401 min
Delta R.T.: -0.004 min
Response: 25453860
Conc: 454.42 ng/ml




#33 AR-1260-3

R.T.: 7.914 min
Delta R.T.: -0.004 min
Response: 39362805
Conc: 464.33 ng/ml



#33 AR-1260-3


R.T.: 6.554 min
Delta R.T.: -0.005 min
Response: 23770176
Conc: 446.70 ng/ml



#34 AR-1260-4


R.T.: 8.143 min
Delta R.T.: -0.004 min
Response: 46316651
Conc: 454.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.026 min
Delta R.T.: -0.004 min
Response: 18887483
Conc: 468.66 ng/ml



#35 AR-1260-5

R.T.: 8.471 min
Delta R.T.: -0.004 min
Response: 85465205
Conc: 470.24 ng/ml



#35 AR-1260-5

R.T.: 7.266 min
Delta R.T.: -0.005 min
Response: 44845563
Conc: 478.07 ng/ml