

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061322\
 Data File : P0087117.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2022 15:28
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 13 15:54:11 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.639	116.5E6	55635488	49.866	49.413
2) SA Decachlor...	10.353	8.651	89946109	43076939	48.875	54.788
Target Compounds						
21) L5 AR-1248-1	5.675	4.729	22554577	9095709	464.881	459.683
22) L5 AR-1248-2	5.952	4.967	33430328	12340548	469.263	454.182
23) L5 AR-1248-3	6.159	5.009	36296341	12997039	475.482	454.788
24) L5 AR-1248-4	6.564	5.180	39375464	14993302	472.948	455.682
25) L5 AR-1248-5	6.604	5.573	37950742	14846767	472.911	464.483



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061322\
Data File : P0087117.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2022 15:28
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 13 15:54:11 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.002 min
Response: 116537185
Conc: 49.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 55635488
Conc: 49.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.353 min
Delta R.T.: -0.003 min
Response: 89946109
Conc: 48.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.651 min
Delta R.T.: -0.006 min
Response: 43076939
Conc: 54.79 ng/ml



#21 AR-1248-1


R.T.: 5.675 min
Delta R.T.: -0.002 min
Response: 22554577
Conc: 464.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.729 min
Delta R.T.: -0.002 min
Response: 9095709
Conc: 459.68 ng/ml




#22 AR-1248-2

R.T.: 5.952 min
Delta R.T.: -0.002 min
Response: 33430328
Conc: 469.26 ng/ml



#22 AR-1248-2


R.T.: 4.967 min
Delta R.T.: -0.003 min
Response: 12340548
Conc: 454.18 ng/ml



#23 AR-1248-3


R.T.: 6.159 min
Delta R.T.: -0.001 min
Response: 36296341
Conc: 475.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 5.009 min
Delta R.T.: -0.002 min
Response: 12997039
Conc: 454.79 ng/ml




#24 AR-1248-4

R.T.: 6.564 min
Delta R.T.: -0.002 min
Response: 39375464
Conc: 472.95 ng/ml



#24 AR-1248-4


R.T.: 5.180 min
Delta R.T.: -0.002 min
Response: 14993302
Conc: 455.68 ng/ml



#25 AR-1248-5

R.T.: 6.604 min
Delta R.T.: -0.003 min
Response: 37950742
Conc: 472.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.573 min
Delta R.T.: -0.002 min
Response: 14846767
Conc: 464.48 ng/ml