

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061323\
 Data File : P0095632.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2023 10:18
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 02:02:56 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 10 06:51:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.405	3.623	174.4E6	73672587	40.680	40.526
2)	SA Decachlor...	10.241	8.677	121.9E6	55654647	48.903	49.972
Target Compounds							
16)	L4 AR-1242-1	5.587	4.716	41266894	16079842	492.699	504.309
17)	L4 AR-1242-2	5.609	4.735	59111967	22892918	478.870	512.764
18)	L4 AR-1242-3	5.672	4.912	38920324	13620948	491.397	530.635
19)	L4 AR-1242-4	5.771	4.997	30555596	14939083	482.573	514.069
20)	L4 AR-1242-5	6.515	5.524	30007353	15914116	487.762	547.002



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061323\
Data File : PO095632.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2023 10:18
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:02:56 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 10 06:51:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.405 min
Delta R.T.: 0.002 min
Response: 174360814
Conc: 40.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: 0.003 min
Response: 73672587
Conc: 40.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.241 min
Delta R.T.: 0.006 min
Response: 121879017
Conc: 48.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: 0.002 min
Response: 55654647
Conc: 49.97 ng/ml



#16 AR-1242-1


R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 41266894
Conc: 492.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.716 min
Delta R.T.: 0.000 min
Response: 16079842
Conc: 504.31 ng/ml




#17 AR-1242-2

R.T.: 5.609 min
Delta R.T.: 0.000 min
Response: 59111967
Conc: 478.87 ng/ml



#17 AR-1242-2


R.T.: 4.735 min
Delta R.T.: 0.000 min
Response: 22892918
Conc: 512.76 ng/ml



#18 AR-1242-3


R.T.: 5.672 min
Delta R.T.: 0.000 min
Response: 38920324
Conc: 491.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.912 min
Delta R.T.: 0.000 min
Response: 13620948
Conc: 530.63 ng/ml




#19 AR-1242-4

R.T.: 5.771 min
Delta R.T.: 0.000 min
Response: 30555596
Conc: 482.57 ng/ml



#19 AR-1242-4


R.T.: 4.997 min
Delta R.T.: 0.000 min
Response: 14939083
Conc: 514.07 ng/ml



#20 AR-1242-5

R.T.: 6.515 min
Delta R.T.: 0.000 min
Response: 30007353
Conc: 487.76 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.524 min
Delta R.T.: 0.000 min
Response: 15914116
Conc: 547.00 ng/ml