

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061323\
 Data File : P0095649.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Jun 2023 20:44
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/14/2023
 Supervised By :Ankita Jodhani 06/14/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 02:13:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 10 06:51:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.404	3.622	182.3E6	75528750	42.542	41.547
2) SA Decachlor...	10.239	8.676	133.2E6	62339694	53.438	55.975
Target Compounds						
26) L6 AR-1254-1	6.449	5.523	57731747	35720062	466.785	514.393m
27) L6 AR-1254-2	6.669	5.672	91649851	32371730	512.556	507.487m
28) L6 AR-1254-3	7.039	6.079	90691079	49982307	523.063	545.221
29) L6 AR-1254-4	7.327	6.309	62289157	26673741	543.183	625.349
30) L6 AR-1254-5	7.750	6.730	69714881	43075095	505.505	554.055

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061323\
Data File : PO095649.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Jun 2023 20:44
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/14/2023
Supervised By :Ankita Jodhani 06/14/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 02:13:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 10 06:51:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.404 min
Delta R.T.: 0.001 min
Response: 182345604
Conc: 42.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/14/2023
Supervised By :Ankita Jodhani 06/14/2023




#1 Tetrachloro-m-xylene

R.T.: 3.622 min
Delta R.T.: 0.002 min
Response: 75528750
Conc: 41.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.239 min
Delta R.T.: 0.004 min
Response: 133181467
Conc: 53.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.676 min
Delta R.T.: 0.000 min
Response: 62339694
Conc: 55.97 ng/ml




#26 AR-1254-1

R.T.: 6.449 min
Delta R.T.: 0.000 min
Response: 57731747
Conc: 466.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/14/2023
Supervised By :Ankita Jodhani 06/14/2023




#26 AR-1254-1

R.T.: 5.523 min
Delta R.T.: 0.000 min
Response: 35720062
Conc: 514.39 ng/ml m




#27 AR-1254-2

R.T.: 6.669 min
Delta R.T.: -0.001 min
Response: 91649851
Conc: 512.56 ng/ml



#27 AR-1254-2

R.T.: 5.672 min
Delta R.T.: 0.000 min
Response: 32371730
Conc: 507.49 ng/ml m




#28 AR-1254-3

R.T.: 7.039 min
Delta R.T.: -0.001 min
Response: 90691079
Conc: 523.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/14/2023
Supervised By :Ankita Jodhani 06/14/2023




#28 AR-1254-3

R.T.: 6.079 min
Delta R.T.: 0.000 min
Response: 49982307
Conc: 545.22 ng/ml




#29 AR-1254-4

R.T.: 7.327 min
Delta R.T.: -0.001 min
Response: 62289157
Conc: 543.18 ng/ml



#29 AR-1254-4

R.T.: 6.309 min
Delta R.T.: 0.000 min
Response: 26673741
Conc: 625.35 ng/ml



#30 AR-1254-5

R.T.: 7.750 min
Delta R.T.: -0.001 min
Response: 69714881
Conc: 505.51 ng/ml

Instrument :

ECD_O


ClientSampleId :

AR1254CCC500

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 06/14/2023
Supervised By :Ankita Jodhani 06/14/2023



#30 AR-1254-5

R.T.: 6.730 min
Delta R.T.: 0.000 min
Response: 43075095
Conc: 554.05 ng/ml