

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061324\
 Data File : P0104324.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2024 00:57
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 01:39:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 12 05:42:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.402	3.546	428.8E6	111.4E6	51.384	51.682
2) SA Decachlor...	10.103	8.458	301.9E6	163.4E6	51.533	51.048
Target Compounds						
16) L4 AR-1242-1	5.563	4.605	120.5E6	29652375	504.803	502.052
17) L4 AR-1242-2	5.586	4.623	166.5E6	45351124	504.492	516.052
18) L4 AR-1242-3	5.647	4.795	105.2E6	25836606	498.355	520.007
19) L4 AR-1242-4	5.746	4.878	87778879	22296451	498.851	516.859
20) L4 AR-1242-5	6.479	5.392	87300377	29629599	499.182	523.649


(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061324\
Data File : PO104324.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2024 00:57
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 01:39:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061124.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 12 05:42:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.402 min
Delta R.T.: 0.000 min
Response: 428828209
Conc: 51.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: 0.000 min
Response: 111425893
Conc: 51.68 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.103 min
Delta R.T.: -0.001 min
Response: 301878882
Conc: 51.53 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.458 min
Delta R.T.: -0.002 min
Response: 163403174
Conc: 51.05 ng/ml



#16 AR-1242-1


R.T.: 5.563 min
Delta R.T.: 0.000 min
Response: 120486841
Conc: 504.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.605 min
Delta R.T.: 0.000 min
Response: 29652375
Conc: 502.05 ng/ml




#17 AR-1242-2

R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 166482802
Conc: 504.49 ng/ml



#17 AR-1242-2


R.T.: 4.623 min
Delta R.T.: 0.000 min
Response: 45351124
Conc: 516.05 ng/ml



#18 AR-1242-3


R.T.: 5.647 min
Delta R.T.: 0.000 min
Response: 105199548
Conc: 498.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.795 min
Delta R.T.: 0.000 min
Response: 25836606
Conc: 520.01 ng/ml




#19 AR-1242-4

R.T.: 5.746 min
Delta R.T.: 0.000 min
Response: 87778879
Conc: 498.85 ng/ml



#19 AR-1242-4


R.T.: 4.878 min
Delta R.T.: -0.002 min
Response: 22296451
Conc: 516.86 ng/ml



#20 AR-1242-5

R.T.: 6.479 min
Delta R.T.: -0.001 min
Response: 87300377
Conc: 499.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.392 min
Delta R.T.: 0.000 min
Response: 29629599
Conc: 523.65 ng/ml