

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061418\
 Data File : P0045792.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jun 2018 22:04
 Operator : SM/UA
 Sample : J3271-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 TP-7A-WC-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 08:25:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 15 08:20:04 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.390	3.624	4416346	2553533	22.510	27.247
2) SA Decachlor...	10.054	8.605	2190588	856332	10.868	10.841
Target Compounds						
3) L1 AR-1016-1	5.268	4.496	1249342	131373	239.435	48.947 #
4) L1 AR-1016-2	5.606	4.940	531306	180770	79.036	83.538
5) L1 AR-1016-3	5.711	5.009	1407133	227352	251.477	88.462 #
6) L1 AR-1016-4	6.192	5.180	464547	37641	78.788	12.795 #
7) L1 AR-1016-5	6.407	5.512	1696818	1927335	314.404	700.568 #
8) L2 AR-1221-1	0.000	3.831	0	13101	N.D.	9.057 #
9) L2 AR-1221-2	4.692	3.923	82236	147745	41.481	141.033 #
10) L2 AR-1221-3	4.780	3.980	88101	175779	14.395	55.261 #
11) L3 AR-1232-1	5.077	3.980	147792	175779	62.519	82.443 #
12) L3 AR-1232-2	5.527	4.711	231274	104409	67.934	64.676
13) L3 AR-1232-3	5.576	4.725	179694	225403	36.810	94.904 #
14) L3 AR-1232-4	5.606	4.776	531306	296173	166.300	180.978
15) L3 AR-1232-5	5.711	4.895	1407133	213133	541.836	171.382 #
16) L4 AR-1242-1	5.077	4.496	147792	131373	37.746	64.795 #
17) L4 AR-1242-2	5.576	4.725	179694	225403	22.754	60.076 #
18) L4 AR-1242-3	5.711	4.895	1407133	213133	332.997	104.724 #
19) L4 AR-1242-4	6.037	5.009	349275	227352	83.386	119.372 #
20) L4 AR-1242-5	6.192	5.180	464547	37641	97.285	18.179 #
21) L5 AR-1248-1	6.037	5.180	349275	37641	48.911	10.379 #
22) L5 AR-1248-2	6.192	5.299	464547	220192	72.894	67.181
23) L5 AR-1248-3	6.407	5.512	1696818	1927335	201.279	367.682 #
24) L5 AR-1248-4	6.478	5.512	230138	1927335	28.095	498.178 #
25) L5 AR-1248-5	6.678	6.070	4895457	1415536	880.127	532.230 #
26) L6 AR-1254-1	6.612	5.512	2586115	1927335	189.216	321.669 #
27) L6 AR-1254-2	6.880	5.647	6776856	570610	798.562	107.881 #
28) L6 AR-1254-3	6.987	6.070	1822677	1415536	122.146	165.931 #
29) L6 AR-1254-4	7.262	6.271	97980	303393	8.598	53.349 #
30) L6 AR-1254-5	7.524	6.578	197680	152087	23.086	37.858 #
31) L7 AR-1260-1	7.142	6.184	350175	627305	32.938	113.069 #
32) L7 AR-1260-2	7.383	6.365	3482353	666272	276.556	95.973 #
33) L7 AR-1260-3	7.679	6.693	793439	260341	52.664	35.961 #
34) L7 AR-1260-4	8.291	7.229	449375	330279	21.497	29.600 #
35) L7 AR-1260-5	8.608	7.576	322621	177339	23.536	22.219
36) L8 AR-1262-1	7.142	6.365	350175	666272	35.141	111.160 #
37) L8 AR-1262-2	7.383	6.519	3482353	390986	299.023	67.252 #
38) L8 AR-1262-3	7.679	6.729	793439	129213	73.965	16.198 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0061418\
 Data File : P0045792.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jun 2018 22:04
 Operator : SM/UA
 Sample : J3271-02
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 TP-7A-WC-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 08:25:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 15 08:20:04 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

39)	L8 AR-1262-4	7.978	6.990	477868	96728	32.330	13.296 #
40)	L8 AR-1262-5	8.291	7.229	449375	330279	16.650	24.296 #
41)	L9 AR-1268-1	8.608	7.517	322621	51819	11.962	4.054 #
42)	L9 AR-1268-2	8.685	7.576	118101	177339	4.655	15.322 #
43)	L9 AR-1268-3	8.881	7.787	268419	45898	12.088	4.678 #
44)	L9 AR-1268-4	9.322	8.065	355346	79153	36.948	17.884 #
45)	L9 AR-1268-5	9.719	8.353	75656	60994	1.139	2.418 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061418\
Data File : P0045792.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jun 2018 22:04
Operator : SM/UA
Sample : J3271-02
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 08:25:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061418.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 15 08:20:04 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.390 min
Delta R.T.: 0.000 min
Response: 4416346
Conc: 22.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 2553533
Conc: 27.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.054 min
Delta R.T.: 0.000 min
Response: 2190588
Conc: 10.87 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.605 min
Delta R.T.: -0.003 min
Response: 856332
Conc: 10.84 ng/ml



#3 AR-1016-1


R.T.: 5.268 min
Delta R.T.: -0.015 min
Response: 1249342
Conc: 239.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#3 AR-1016-1

R.T.: 4.496 min
Delta R.T.: 0.009 min
Response: 131373
Conc: 48.95 ng/ml




#4 AR-1016-2

R.T.: 5.606 min
Delta R.T.: -0.027 min
Response: 531306
Conc: 79.04 ng/ml



#4 AR-1016-2


R.T.: 4.940 min
Delta R.T.: -0.001 min
Response: 180770
Conc: 83.54 ng/ml



#5 AR-1016-3


R.T.: 5.711 min
Delta R.T.: -0.019 min
Response: 1407133
Conc: 251.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#5 AR-1016-3

R.T.: 5.009 min
Delta R.T.: 0.027 min
Response: 227352
Conc: 88.46 ng/ml




#6 AR-1016-4

R.T.: 6.192 min
Delta R.T.: 0.029 min
Response: 464547
Conc: 78.79 ng/ml



#6 AR-1016-4


R.T.: 5.180 min
Delta R.T.: 0.026 min
Response: 37641
Conc: 12.79 ng/ml



#7 AR-1016-5


R.T.: 6.407 min
Delta R.T.: 0.009 min
Response: 1696818
Conc: 314.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#7 AR-1016-5

R.T.: 5.512 min
Delta R.T.: 0.009 min
Response: 1927335
Conc: 700.57 ng/ml




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T.: 4.595 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 3.831 min
Delta R.T.: -0.007 min
Response: 13101
Conc: 9.06 ng/ml



#9 AR-1221-2


R.T.: 4.692 min
Delta R.T.: 0.011 min
Response: 82236
Conc: 41.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#9 AR-1221-2

R.T.: 3.923 min
Delta R.T.: 0.000 min
Response: 147745
Conc: 141.03 ng/ml




#10 AR-1221-3

R.T.: 4.780 min
Delta R.T.: 0.023 min
Response: 88101
Conc: 14.39 ng/ml



#10 AR-1221-3


R.T.: 3.980 min
Delta R.T.: -0.019 min
Response: 175779
Conc: 55.26 ng/ml



#11 AR-1232-1


R.T.: 5.077 min
Delta R.T.: -0.014 min
Response: 147792
Conc: 62.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#11 AR-1232-1

R.T.: 3.980 min
Delta R.T.: -0.019 min
Response: 175779
Conc: 82.44 ng/ml




#12 AR-1232-2

R.T.: 5.527 min
Delta R.T.: -0.021 min
Response: 231274
Conc: 67.93 ng/ml



#12 AR-1232-2


R.T.: 4.711 min
Delta R.T.: 0.006 min
Response: 104409
Conc: 64.68 ng/ml



#13 AR-1232-3


R.T.: 5.576 min
Delta R.T.: 0.005 min
Response: 179694
Conc: 36.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#13 AR-1232-3

R.T.: 4.725 min
Delta R.T.: 0.000 min
Response: 225403
Conc: 94.90 ng/ml




#14 AR-1232-4

R.T.: 5.606 min
Delta R.T.: -0.027 min
Response: 531306
Conc: 166.30 ng/ml



#14 AR-1232-4


R.T.: 4.776 min
Delta R.T.: -0.003 min
Response: 296173
Conc: 180.98 ng/ml



#15 AR-1232-5


R.T.: 5.711 min
Delta R.T.: -0.018 min
Response: 1407133
Conc: 541.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#15 AR-1232-5

R.T.: 4.895 min
Delta R.T.: -0.004 min
Response: 213133
Conc: 171.38 ng/ml




#16 AR-1242-1

R.T.: 5.077 min
Delta R.T.: -0.013 min
Response: 147792
Conc: 37.75 ng/ml



#16 AR-1242-1


R.T.: 4.496 min
Delta R.T.: 0.009 min
Response: 131373
Conc: 64.80 ng/ml



#17 AR-1242-2


R.T.: 5.576 min
Delta R.T.: 0.006 min
Response: 179694
Conc: 22.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#17 AR-1242-2

R.T.: 4.725 min
Delta R.T.: 0.000 min
Response: 225403
Conc: 60.08 ng/ml




#18 AR-1242-3

R.T.: 5.711 min
Delta R.T.: -0.018 min
Response: 1407133
Conc: 333.00 ng/ml



#18 AR-1242-3


R.T.: 4.895 min
Delta R.T.: -0.004 min
Response: 213133
Conc: 104.72 ng/ml



#19 AR-1242-4


R.T.: 6.037 min
Delta R.T.: 0.015 min
Response: 349275
Conc: 83.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#19 AR-1242-4

R.T.: 5.009 min
Delta R.T.: 0.027 min
Response: 227352
Conc: 119.37 ng/ml




#20 AR-1242-5

R.T.: 6.192 min
Delta R.T.: 0.030 min
Response: 464547
Conc: 97.29 ng/ml



#20 AR-1242-5


R.T.: 5.180 min
Delta R.T.: 0.027 min
Response: 37641
Conc: 18.18 ng/ml



#21 AR-1248-1


R.T.: 6.037 min
Delta R.T.: 0.014 min
Response: 349275
Conc: 48.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#21 AR-1248-1

R.T.: 5.180 min
Delta R.T.: 0.027 min
Response: 37641
Conc: 10.38 ng/ml




#22 AR-1248-2

R.T.: 6.192 min
Delta R.T.: 0.030 min
Response: 464547
Conc: 72.89 ng/ml



#22 AR-1248-2


R.T.: 5.299 min
Delta R.T.: -0.001 min
Response: 220192
Conc: 67.18 ng/ml



#23 AR-1248-3


R.T.: 6.407 min
Delta R.T.: -0.016 min
Response: 1696818
Conc: 201.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#23 AR-1248-3

R.T.: 5.512 min
Delta R.T.: 0.010 min
Response: 1927335
Conc: 367.68 ng/ml




#24 AR-1248-4

R.T.: 6.478 min
Delta R.T.: 0.016 min
Response: 230138
Conc: 28.09 ng/ml



#24 AR-1248-4


R.T.: 5.512 min
Delta R.T.: -0.031 min
Response: 1927335
Conc: 498.18 ng/ml



#25 AR-1248-5


R.T.: 6.678 min
Delta R.T.: 0.021 min
Response: 4895457
Conc: 880.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#25 AR-1248-5

R.T.: 6.070 min
Delta R.T.: 0.018 min
Response: 1415536
Conc: 532.23 ng/ml




#26 AR-1254-1

R.T.: 6.612 min
Delta R.T.: -0.004 min
Response: 2586115
Conc: 189.22 ng/ml



#26 AR-1254-1


R.T.: 5.512 min
Delta R.T.: 0.009 min
Response: 1927335
Conc: 321.67 ng/ml



#27 AR-1254-2


R.T.: 6.880 min
Delta R.T.: -0.014 min
Response: 6776856
Conc: 798.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#27 AR-1254-2

R.T.: 5.647 min
Delta R.T.: -0.002 min
Response: 570610
Conc: 107.88 ng/ml




#28 AR-1254-3

R.T.: 6.987 min
Delta R.T.: 0.008 min
Response: 1822677
Conc: 122.15 ng/ml



#28 AR-1254-3


R.T.: 6.070 min
Delta R.T.: 0.017 min
Response: 1415536
Conc: 165.93 ng/ml



#29 AR-1254-4


R.T.: 7.262 min
Delta R.T.: 0.000 min
Response: 97980
Conc: 8.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#29 AR-1254-4

R.T.: 6.271 min
Delta R.T.: -0.008 min
Response: 303393
Conc: 53.35 ng/ml




#30 AR-1254-5

R.T.: 7.524 min
Delta R.T.: -0.008 min
Response: 197680
Conc: 23.09 ng/ml



#30 AR-1254-5


R.T.: 6.578 min
Delta R.T.: -0.015 min
Response: 152087
Conc: 37.86 ng/ml



#31 AR-1260-1


R.T.: 7.142 min
Delta R.T.: -0.001 min
Response: 350175
Conc: 32.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#31 AR-1260-1

R.T.: 6.184 min
Delta R.T.: 0.002 min
Response: 627305
Conc: 113.07 ng/ml




#32 AR-1260-2

R.T.: 7.383 min
Delta R.T.: -0.016 min
Response: 3482353
Conc: 276.56 ng/ml



#32 AR-1260-2


R.T.: 6.365 min
Delta R.T.: -0.003 min
Response: 666272
Conc: 95.97 ng/ml



#33 AR-1260-3


R.T.: 7.679 min
Delta R.T.: 0.001 min
Response: 793439
Conc: 52.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#33 AR-1260-3

R.T.: 6.693 min
Delta R.T.: -0.003 min
Response: 260341
Conc: 35.96 ng/ml




#34 AR-1260-4

R.T.: 8.291 min
Delta R.T.: -0.001 min
Response: 449375
Conc: 21.50 ng/ml



#34 AR-1260-4


R.T.: 7.229 min
Delta R.T.: -0.003 min
Response: 330279
Conc: 29.60 ng/ml



#35 AR-1260-5


R.T.: 8.608 min
Delta R.T.: 0.000 min
Response: 322621
Conc: 23.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#35 AR-1260-5

R.T.: 7.576 min
Delta R.T.: -0.003 min
Response: 177339
Conc: 22.22 ng/ml




#36 AR-1262-1

R.T.: 7.142 min
Delta R.T.: 0.000 min
Response: 350175
Conc: 35.14 ng/ml



#36 AR-1262-1


R.T.: 6.365 min
Delta R.T.: -0.002 min
Response: 666272
Conc: 111.16 ng/ml



#37 AR-1262-2


R.T.: 7.383 min
Delta R.T.: -0.015 min
Response: 3482353
Conc: 299.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#37 AR-1262-2

R.T.: 6.519 min
Delta R.T.: -0.005 min
Response: 390986
Conc: 67.25 ng/ml




#38 AR-1262-3

R.T.: 7.679 min
Delta R.T.: 0.003 min
Response: 793439
Conc: 73.97 ng/ml



#38 AR-1262-3


R.T.: 6.729 min
Delta R.T.: -0.004 min
Response: 129213
Conc: 16.20 ng/ml



#39 AR-1262-4


R.T.: 7.978 min
Delta R.T.: -0.002 min
Response: 477868
Conc: 32.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#39 AR-1262-4

R.T.: 6.990 min
Delta R.T.: -0.002 min
Response: 96728
Conc: 13.30 ng/ml




#40 AR-1262-5

R.T.: 8.291 min
Delta R.T.: -0.001 min
Response: 449375
Conc: 16.65 ng/ml



#40 AR-1262-5


R.T.: 7.229 min
Delta R.T.: -0.003 min
Response: 330279
Conc: 24.30 ng/ml



#41 AR-1268-1


R.T.: 8.608 min
Delta R.T.: 0.007 min
Response: 322621
Conc: 11.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#41 AR-1268-1

R.T.: 7.517 min
Delta R.T.: 0.000 min
Response: 51819
Conc: 4.05 ng/ml




#42 AR-1268-2

R.T.: 8.685 min
Delta R.T.: -0.008 min
Response: 118101
Conc: 4.66 ng/ml



#42 AR-1268-2


R.T.: 7.576 min
Delta R.T.: -0.004 min
Response: 177339
Conc: 15.32 ng/ml



#43 AR-1268-3


R.T.: 8.881 min
Delta R.T.: -0.037 min
Response: 268419
Conc: 12.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1




#43 AR-1268-3

R.T.: 7.787 min
Delta R.T.: -0.002 min
Response: 45898
Conc: 4.68 ng/ml




#44 AR-1268-4

R.T.: 9.322 min
Delta R.T.: -0.003 min
Response: 355346
Conc: 36.95 ng/ml



#44 AR-1268-4


R.T.: 8.065 min
Delta R.T.: -0.006 min
Response: 79153
Conc: 17.88 ng/ml



#45 AR-1268-5

R.T.: 9.719 min
Delta R.T.: -0.012 min
Response: 75656
Conc: 1.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
TP-7A-WC-1



#45 AR-1268-5

R.T.: 8.353 min
Delta R.T.: -0.004 min
Response: 60994
Conc: 2.42 ng/ml