

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\P0061418\
 Data File : P0045770.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jun 2018 15:40
 Operator : SM/UA
 Sample : P0061418ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO061418

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 15:55:53 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0061418.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 14 15:46:28 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.391	3.624	10042530	4589259	51.186	48.969
2) SA Decachlor...	10.056	8.607	9803520	3894277	48.639	49.300
Target Compounds						
3) L1 AR-1016-1	5.283	4.486	2811466	1406205	538.814	523.922
4) L1 AR-1016-2	5.633	4.940	3596782	1090575	535.050	503.979
5) L1 AR-1016-3	5.730	4.982	3062265	1323283	547.276	514.884
6) L1 AR-1016-4	6.163	5.153	3199209	1424859	542.594	484.329
7) L1 AR-1016-5	6.398	5.503	2760800	1457324	511.550	529.724
31) L7 AR-1260-1	7.144	6.182	5593365	2856249	526.115	514.825
32) L7 AR-1260-2	7.399	6.367	6602145	3492180	524.318	503.033
33) L7 AR-1260-3	7.679	6.694	7659940	3697647	508.426	510.761
34) L7 AR-1260-4	8.293	7.231	10537005	5779135	504.068	517.935
35) L7 AR-1260-5	8.607	7.578	6871241	4049136	501.272	507.329



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD 0\Data\PO061418\
Data File : PO045770.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jun 2018 15:40
Operator : SM/UA
Sample : PO061418ICV500
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO061418

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 15:55:53 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\PO061418.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 14 15:46:28 2018
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Ini. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.391 min
Delta R.T.: -0.003 min
Response: 10042530
Conc: 51.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO061418




#1 Tetrachloro-m-xylene

R.T.: 3.624 min
Delta R.T.: 0.000 min
Response: 4589259
Conc: 48.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.056 min
Delta R.T.: -0.003 min
Response: 9803520
Conc: 48.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.607 min
Delta R.T.: -0.003 min
Response: 3894277
Conc: 49.30 ng/ml



#3 AR-1016-1


R.T.: 5.283 min
Delta R.T.: -0.002 min
Response: 2811466
Conc: 538.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO061418




#3 AR-1016-1

R.T.: 4.486 min
Delta R.T.: 0.000 min
Response: 1406205
Conc: 523.92 ng/ml




#4 AR-1016-2

R.T.: 5.633 min
Delta R.T.: -0.001 min
Response: 3596782
Conc: 535.05 ng/ml



#4 AR-1016-2


R.T.: 4.940 min
Delta R.T.: -0.003 min
Response: 1090575
Conc: 503.98 ng/ml



#5 AR-1016-3


R.T.: 5.730 min
Delta R.T.: -0.002 min
Response: 3062265
Conc: 547.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO061418




#5 AR-1016-3

R.T.: 4.982 min
Delta R.T.: 0.000 min
Response: 1323283
Conc: 514.88 ng/ml




#6 AR-1016-4

R.T.: 6.163 min
Delta R.T.: -0.002 min
Response: 3199209
Conc: 542.59 ng/ml



#6 AR-1016-4


R.T.: 5.153 min
Delta R.T.: -0.002 min
Response: 1424859
Conc: 484.33 ng/ml



#7 AR-1016-5


R.T.: 6.398 min
Delta R.T.: -0.002 min
Response: 2760800
Conc: 511.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO061418




#7 AR-1016-5

R.T.: 5.503 min
Delta R.T.: 0.002 min
Response: 1457324
Conc: 529.72 ng/ml




#31 AR-1260-1

R.T.: 7.144 min
Delta R.T.: -0.001 min
Response: 5593365
Conc: 526.12 ng/ml



#31 AR-1260-1


R.T.: 6.182 min
Delta R.T.: -0.001 min
Response: 2856249
Conc: 514.82 ng/ml



#32 AR-1260-2


R.T.: 7.399 min
Delta R.T.: -0.003 min
Response: 6602145
Conc: 524.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO061418




#32 AR-1260-2

R.T.: 6.367 min
Delta R.T.: 0.000 min
Response: 3492180
Conc: 503.03 ng/ml




#33 AR-1260-3

R.T.: 7.679 min
Delta R.T.: -0.001 min
Response: 7659940
Conc: 508.43 ng/ml



#33 AR-1260-3


R.T.: 6.694 min
Delta R.T.: -0.002 min
Response: 3697647
Conc: 510.76 ng/ml



#34 AR-1260-4


R.T.: 8.293 min
Delta R.T.: -0.001 min
Response: 10537005
Conc: 504.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO061418




#34 AR-1260-4

R.T.: 7.231 min
Delta R.T.: 0.000 min
Response: 5779135
Conc: 517.93 ng/ml



#35 AR-1260-5

R.T.: 8.607 min
Delta R.T.: -0.003 min
Response: 6871241
Conc: 501.27 ng/ml



#35 AR-1260-5

R.T.: 7.578 min
Delta R.T.: -0.002 min
Response: 4049136
Conc: 507.33 ng/ml