

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061421\
 Data File : P0078775.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jun 2021 18:09
 Operator : DD\AJ
 Sample : M2677-19
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 01:42:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.330	3.606	1682066	707484	19.867	20.437
2) SA Decachlor...	9.938	8.789	1230657	659038	20.971	21.196
Target Compounds						
3) L1 AR-1016-1	5.618	0.000	11293	0	4.423	N.D. #
4) L1 AR-1016-2	5.618f	0.000	11293	0	2.898	N.D. #
8) L2 AR-1221-1	4.576	0.000	66385	0	73.770	N.D. #
9) L2 AR-1221-2	4.672	0.000	16894	0	25.463	N.D. #
12) L3 AR-1232-2	5.360f	0.000	77811	0	94.949	N.D. #
16) L4 AR-1242-1	5.618	0.000	11293	0	5.733	N.D. #
20) L4 AR-1242-5	6.581	0.000	1107	0	0.736	N.D. #
21) L5 AR-1248-1	5.618	0.000	11293	0	7.352	N.D. #
24) L5 AR-1248-4	6.549	0.000	5690	0	2.173	N.D. #
25) L5 AR-1248-5	6.581	0.000	1107	0	0.438	N.D. #
26) L6 AR-1254-1	6.532	0.000	20664	0	7.538	N.D. #
28) L6 AR-1254-3	7.138	0.000	24184	0	5.555	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061421\
Data File : P0078775.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jun 2021 18:09
Operator : DD\AJ
Sample : M2677-19
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 01:42:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 14 05:33:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.001 min
Response: 1682066
Conc: 19.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 707484
Conc: 20.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.938 min
Delta R.T.: 0.000 min
Response: 1230657
Conc: 20.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.789 min
Delta R.T.: -0.001 min
Response: 659038
Conc: 21.20 ng/ml



#3 AR-1016-1


R.T.: 5.618 min
Delta R.T.: -0.008 min
Response: 11293
Conc: 4.42 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.834 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.618 min
Delta R.T.: -0.030 min
Response: 11293
Conc: 2.90 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.853 min
Response: 0
Conc: N.D.



#8 AR-1221-1


R.T.: 4.576 min
Delta R.T.: -0.006 min
Response: 66385
Conc: 73.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 3.859 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 4.672 min
Delta R.T.: -0.004 min
Response: 16894
Conc: 25.46 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 3.956 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.360 min
Delta R.T.: 0.025 min
Response: 77811
Conc: 94.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 4.852 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.618 min
Delta R.T.: -0.008 min
Response: 11293
Conc: 5.73 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.834 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.581 min
Delta R.T.: -0.008 min
Response: 1107
Conc: 0.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 0.000 min
Exp R.T. : 5.690 min
Response: 0
Conc: N.D.




#21 AR-1248-1

R.T.: 5.618 min
Delta R.T.: -0.009 min
Response: 11293
Conc: 7.35 ng/ml



#21 AR-1248-1


R.T.: 0.000 min
Exp R.T. : 4.833 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.549 min
Delta R.T.: -0.002 min
Response: 5690
Conc: 2.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.316 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: -0.008 min
Response: 1107
Conc: 0.44 ng/ml



#25 AR-1248-5


R.T.: 0.000 min
Exp R.T. : 5.732 min
Response: 0
Conc: N.D.



#26 AR-1254-1


R.T.: 6.532 min
Delta R.T.: 0.011 min
Response: 20664
Conc: 7.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#26 AR-1254-1

R.T.: 0.000 min
Exp R.T. : 5.689 min
Response: 0
Conc: N.D.



#28 AR-1254-3

R.T.: 7.138 min
Delta R.T.: 0.015 min
Response: 24184
Conc: 5.55 ng/ml



#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.260 min
Response: 0
Conc: N.D.