

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061421\
 Data File : P0078764.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 14 Jun 2021 14:59
 Operator : DD\AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 01:37:12 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.330	3.606	3927495	1574091	46.388	45.470
2) SA Decachlor...	9.939	8.790	2754843	1429963	46.944	45.991
Target Compounds						
26) L6 AR-1254-1	6.519	5.688	1336031	829145	487.401	469.302
27) L6 AR-1254-2	6.747	5.847	2101894	747992	496.947	461.664
28) L6 AR-1254-3	7.122	6.259	2173810	1133917	499.287	466.420
29) L6 AR-1254-4	7.416	6.497	1560248	704820	506.838	470.718
30) L6 AR-1254-5	7.838	6.919	1703864	1090645	494.309	475.118



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061421\
Data File : PO078764.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 14 Jun 2021 14:59
Operator : DD\AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 01:37:12 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061321.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 14 05:33:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: -0.001 min
Response: 3927495
Conc: 46.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 1574091
Conc: 45.47 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.939 min
Delta R.T.: 0.001 min
Response: 2754843
Conc: 46.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.790 min
Delta R.T.: 0.000 min
Response: 1429963
Conc: 45.99 ng/ml



#26 AR-1254-1


R.T.: 6.519 min
Delta R.T.: -0.002 min
Response: 1336031
Conc: 487.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.688 min
Delta R.T.: 0.000 min
Response: 829145
Conc: 469.30 ng/ml




#27 AR-1254-2

R.T.: 6.747 min
Delta R.T.: -0.001 min
Response: 2101894
Conc: 496.95 ng/ml



#27 AR-1254-2


R.T.: 5.847 min
Delta R.T.: 0.000 min
Response: 747992
Conc: 461.66 ng/ml



#28 AR-1254-3


R.T.: 7.122 min
Delta R.T.: -0.001 min
Response: 2173810
Conc: 499.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.259 min
Delta R.T.: 0.000 min
Response: 1133917
Conc: 466.42 ng/ml




#29 AR-1254-4

R.T.: 7.416 min
Delta R.T.: -0.002 min
Response: 1560248
Conc: 506.84 ng/ml



#29 AR-1254-4


R.T.: 6.497 min
Delta R.T.: -0.001 min
Response: 704820
Conc: 470.72 ng/ml



#30 AR-1254-5

R.T.: 7.838 min
Delta R.T.: -0.001 min
Response: 1703864
Conc: 494.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.919 min
Delta R.T.: -0.001 min
Response: 1090645
Conc: 475.12 ng/ml