

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061422\
 Data File : P0087146.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2022 12:39
 Operator : YP/AJ
 Sample : N3315-09 10X
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 BC107651-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 06/15/2022
 Supervised By :Ankita Jodhani 06/15/2022

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 13:19:27 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.491	3.639	3266551	1702862	1.398	1.512
2) SA Decachlor...	10.347	8.650	3975679	1178714	2.160m	1.499 #

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061422\
Data File : PO087146.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2022 12:39
Operator : YP/AJ
Sample : N3315-09 10X
Misc :
ALS Vial : 11 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
BC107651-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/15/2022
Supervised By :Ankita Jodhani 06/15/2022

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 13:19:27 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.491 min
Delta R.T.: -0.002 min
Response: 3266551
Conc: 1.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
BC107651-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 06/15/2022
Supervised By :Ankita Jodhani 06/15/2022




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 1702862
Conc: 1.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.347 min
Delta R.T.: -0.009 min
Response: 3975679
Conc: 2.16 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.650 min
Delta R.T.: -0.008 min
Response: 1178714
Conc: 1.50 ng/ml