

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061423\
 Data File : P0095699.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Jun 2023 18:15
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 100 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 14 18:58:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 10 06:51:01 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.403	3.620	183.4E6	76959568	42.780	42.334
2) SA Decachlor...	10.238	8.674	131.5E6	62951809	52.745	56.525
Target Compounds						
3) L1 AR-1016-1	5.584	4.713	56986041	22777234	456.632	477.619
4) L1 AR-1016-2	5.607	4.732	81355598	32695921	445.314	481.691
5) L1 AR-1016-3	5.669	4.910	52367356	18677180	444.958	502.775
6) L1 AR-1016-4	5.769	4.952	41583089	16316156	447.563	459.274
7) L1 AR-1016-5	6.067	5.166	40892886	19735521	440.477	467.062
31) L7 AR-1260-1	7.205	6.208	74801167	36776187	474.731	482.529
32) L7 AR-1260-2	7.464	6.398	84739711	44536271	493.986	556.419
33) L7 AR-1260-3	7.749	6.553	100.7E6	39863472	515.057	514.165
34) L7 AR-1260-4	8.055	7.028	65769397	29353818	515.965	547.233
35) L7 AR-1260-5	8.379	7.272	120.8E6	62571275	546.792	585.628m



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061423\
Data File : P0095699.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Jun 2023 18:15
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 100 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 14 18:58:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060923.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 10 06:51:01 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.403 min
Delta R.T.: 0.000 min
Response: 183363428
Conc: 42.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: 0.000 min
Response: 76959568
Conc: 42.33 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.238 min
Delta R.T.: 0.003 min
Response: 131456082
Conc: 52.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.674 min
Delta R.T.: -0.001 min
Response: 62951809
Conc: 56.52 ng/ml



#3 AR-1016-1


R.T.: 5.584 min
Delta R.T.: 0.000 min
Response: 56986041
Conc: 456.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.713 min
Delta R.T.: -0.001 min
Response: 22777234
Conc: 477.62 ng/ml




#4 AR-1016-2

R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 81355598
Conc: 445.31 ng/ml



#4 AR-1016-2


R.T.: 4.732 min
Delta R.T.: 0.000 min
Response: 32695921
Conc: 481.69 ng/ml



#5 AR-1016-3


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 52367356
Conc: 444.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.910 min
Delta R.T.: 0.000 min
Response: 18677180
Conc: 502.78 ng/ml




#6 AR-1016-4

R.T.: 5.769 min
Delta R.T.: 0.000 min
Response: 41583089
Conc: 447.56 ng/ml



#6 AR-1016-4


R.T.: 4.952 min
Delta R.T.: -0.001 min
Response: 16316156
Conc: 459.27 ng/ml



#7 AR-1016-5


R.T.: 6.067 min
Delta R.T.: 0.000 min
Response: 40892886
Conc: 440.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.166 min
Delta R.T.: 0.000 min
Response: 19735521
Conc: 467.06 ng/ml




#31 AR-1260-1

R.T.: 7.205 min
Delta R.T.: 0.000 min
Response: 74801167
Conc: 474.73 ng/ml



#31 AR-1260-1


R.T.: 6.208 min
Delta R.T.: 0.000 min
Response: 36776187
Conc: 482.53 ng/ml



#32 AR-1260-2


R.T.: 7.464 min
Delta R.T.: 0.000 min
Response: 84739711
Conc: 493.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.398 min
Delta R.T.: -0.002 min
Response: 44536271
Conc: 556.42 ng/ml




#33 AR-1260-3

R.T.: 7.749 min
Delta R.T.: 0.000 min
Response: 100684552
Conc: 515.06 ng/ml



#33 AR-1260-3


R.T.: 6.553 min
Delta R.T.: 0.000 min
Response: 39863472
Conc: 514.16 ng/ml



#34 AR-1260-4


R.T.: 8.055 min
Delta R.T.: 0.000 min
Response: 65769397
Conc: 515.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.028 min
Delta R.T.: -0.001 min
Response: 29353818
Conc: 547.23 ng/ml



#35 AR-1260-5

R.T.: 8.379 min
Delta R.T.: 0.000 min
Response: 120804131
Conc: 546.79 ng/ml



#35 AR-1260-5

R.T.: 7.272 min
Delta R.T.: 0.000 min
Response: 62571275
Conc: 585.63 ng/ml m