

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061521\
 Data File : P0078788.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 15 Jun 2021 9:59
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 15 16:30:53 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.330	3.606	1803419	724610	21.300	20.932
2)	SA Decachlor...	9.941	8.789	1376705	718726	23.460	23.116
Target Compounds							
8)	L2 AR-1221-1	4.576	0.000	27772	0	30.862	N.D. #
9)	L2 AR-1221-2	4.673	0.000	20297	0	30.591	N.D. #
20)	L4 AR-1242-5	6.583	5.713f	4264	100993	2.836	112.889 #
24)	L5 AR-1248-4	6.530f	0.000	282758	0	107.980	N.D. #
25)	L5 AR-1248-5	6.583	5.713f	4264	100993	1.687	91.612 #
26)	L6 AR-1254-1	6.530	5.713f	282758	100993	103.154	57.163 #
28)	L6 AR-1254-3	7.141f	0.000	269159	0	61.821	N.D. #
29)	L6 AR-1254-4	7.417	0.000	11232	0	3.649	N.D. #
30)	L6 AR-1254-5	7.826	0.000	22470	0	6.519	N.D. #
32)	L7 AR-1260-2	0.000	6.562f	0	121362	N.D.	60.221 #
43)	L9 AR-1268-3	9.011f	0.000	23077	0	3.731	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061521\
Data File : P0078788.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 15 Jun 2021 9:59
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 15 16:30:53 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 14 05:33:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.330 min
Delta R.T.: 0.000 min
Response: 1803419
Conc: 21.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.606 min
Delta R.T.: 0.000 min
Response: 724610
Conc: 20.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.941 min
Delta R.T.: 0.003 min
Response: 1376705
Conc: 23.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.789 min
Delta R.T.: -0.001 min
Response: 718726
Conc: 23.12 ng/ml



#8 AR-1221-1


R.T.: 4.576 min
Delta R.T.: -0.005 min
Response: 27772
Conc: 30.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#8 AR-1221-1

R.T.: 0.000 min
Exp R.T. : 3.859 min
Response: 0
Conc: N.D.




#9 AR-1221-2

R.T.: 4.673 min
Delta R.T.: -0.004 min
Response: 20297
Conc: 30.59 ng/ml



#9 AR-1221-2


R.T.: 0.000 min
Exp R.T. : 3.956 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.583 min
Delta R.T.: -0.006 min
Response: 4264
Conc: 2.84 ng/ml

Instrument :
ECD_O
ClientSampleId :




#20 AR-1242-5

R.T.: 5.713 min
Delta R.T.: 0.023 min
Response: 100993
Conc: 112.89 ng/ml




#24 AR-1248-4

R.T.: 6.530 min
Delta R.T.: -0.021 min
Response: 282758
Conc: 107.98 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 5.316 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.583 min
Delta R.T.: -0.005 min
Response: 4264
Conc: 1.69 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.713 min
Delta R.T.: -0.019 min
Response: 100993
Conc: 91.61 ng/ml




#26 AR-1254-1

R.T.: 6.530 min
Delta R.T.: 0.009 min
Response: 282758
Conc: 103.15 ng/ml



#26 AR-1254-1


R.T.: 5.713 min
Delta R.T.: 0.023 min
Response: 100993
Conc: 57.16 ng/ml



#28 AR-1254-3


R.T.: 7.141 min
Delta R.T.: 0.018 min
Response: 269159
Conc: 61.82 ng/ml

Instrument :
ECD_O
ClientSampleId :




#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.260 min
Response: 0
Conc: N.D.




#29 AR-1254-4

R.T.: 7.417 min
Delta R.T.: 0.000 min
Response: 11232
Conc: 3.65 ng/ml



#29 AR-1254-4


R.T.: 0.000 min
Exp R.T. : 6.498 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.826 min
Delta R.T.: -0.013 min
Response: 22470
Conc: 6.52 ng/ml

Instrument :
ECD_O
ClientSampleId :




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T. : 6.921 min
Response: 0
Conc: N.D.




#32 AR-1260-2

R.T.: 0.000 min
Exp R.T. : 7.553 min
Response: 0
Conc: N.D.



#32 AR-1260-2


R.T.: 6.562 min
Delta R.T.: -0.030 min
Response: 121362
Conc: 60.22 ng/ml



#43 AR-1268-3

R.T.: 9.011 min
Delta R.T.: 0.027 min
Response: 23077
Conc: 3.73 ng/ml

Instrument :
ECD_O
ClientSampleId :



#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 8.014 min
Response: 0
Conc: N.D.