

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061522\
 Data File : P0087201.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 15 Jun 2022 19:21
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 16 01:31:37 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.491	3.639	107.5E6	52391576	45.985	46.532
2)	SA Decachlor...	10.348	8.646	83474364	38813123	45.358	49.365
Target Compounds							
16)	L4 AR-1242-1	5.674	4.727	28556234	12479131	444.288	471.483
17)	L4 AR-1242-2	5.696	4.746	40016846	16951762	437.947	470.611
18)	L4 AR-1242-3	5.760	4.921	25746990	9422605	438.650	471.534
19)	L4 AR-1242-4	5.859	5.006	20514458	9098842	440.341	465.507
20)	L4 AR-1242-5	6.602	5.528	21886239	11049301	461.263	473.299



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061522\
Data File : PO087201.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 15 Jun 2022 19:21
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 16 01:31:37 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.491 min
Delta R.T.: -0.002 min
Response: 107466944
Conc: 45.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.639 min
Delta R.T.: -0.002 min
Response: 52391576
Conc: 46.53 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: -0.008 min
Response: 83474364
Conc: 45.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.646 min
Delta R.T.: -0.012 min
Response: 38813123
Conc: 49.36 ng/ml



#16 AR-1242-1


R.T.: 5.674 min
Delta R.T.: -0.003 min
Response: 28556234
Conc: 444.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.727 min
Delta R.T.: -0.005 min
Response: 12479131
Conc: 471.48 ng/ml




#17 AR-1242-2

R.T.: 5.696 min
Delta R.T.: -0.003 min
Response: 40016846
Conc: 437.95 ng/ml



#17 AR-1242-2


R.T.: 4.746 min
Delta R.T.: -0.005 min
Response: 16951762
Conc: 470.61 ng/ml



#18 AR-1242-3


R.T.: 5.760 min
Delta R.T.: -0.002 min
Response: 25746990
Conc: 438.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.921 min
Delta R.T.: -0.006 min
Response: 9422605
Conc: 471.53 ng/ml




#19 AR-1242-4

R.T.: 5.859 min
Delta R.T.: -0.002 min
Response: 20514458
Conc: 440.34 ng/ml



#19 AR-1242-4


R.T.: 5.006 min
Delta R.T.: -0.006 min
Response: 9098842
Conc: 465.51 ng/ml



#20 AR-1242-5

R.T.: 6.602 min
Delta R.T.: -0.003 min
Response: 21886239
Conc: 461.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.528 min
Delta R.T.: -0.006 min
Response: 11049301
Conc: 473.30 ng/ml