

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061623\
 Data File : P0095746.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 16 Jun 2023 23:00
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 17 03:19:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 17 03:19:19 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.428	3.634	115.4E6	66331224	50.000	50.000
2) SA Decachlor...	10.309	8.677	34985248	51454742	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.612	4.725	27833922	20901979	500.000	500.000
4) L1 AR-1016-2	5.635	4.743	41803032	29450972	500.000	500.000
5) L1 AR-1016-3	5.698	4.921	27226052	16030281	500.000	500.000
6) L1 AR-1016-4	5.797	4.963	21103086	13862135	500.000	500.000
7) L1 AR-1016-5	6.097	5.177	16534945	17165322	500.000	500.000
31) L7 AR-1260-1	7.241	6.217	20460052	32471722	500.000	500.000
32) L7 AR-1260-2	7.501	6.406	22486979	37254776	500.000	500.000
33) L7 AR-1260-3	7.787	6.560	27163979	35318401	500.000	500.000
34) L7 AR-1260-4	8.095	7.036	19711076	25439075	500.000	500.000
35) L7 AR-1260-5	8.423	7.278	31932860	57556124	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061623\
Data File : PO095746.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 16 Jun 2023 23:00
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 03:19:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061623.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 17 03:19:19 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: 0.000 min
Response: 115350271
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.634 min
Delta R.T.: 0.000 min
Response: 66331224
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.309 min
Delta R.T.: 0.000 min
Response: 34985248
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.677 min
Delta R.T.: 0.000 min
Response: 51454742
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.612 min
Delta R.T.: 0.000 min
Response: 27833922
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.725 min
Delta R.T.: 0.000 min
Response: 20901979
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.635 min
Delta R.T.: 0.000 min
Response: 41803032
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.743 min
Delta R.T.: 0.000 min
Response: 29450972
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.698 min
Delta R.T.: 0.000 min
Response: 27226052
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.921 min
Delta R.T.: 0.000 min
Response: 16030281
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.797 min
Delta R.T.: 0.000 min
Response: 21103086
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.963 min
Delta R.T.: 0.000 min
Response: 13862135
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.097 min
Delta R.T.: 0.000 min
Response: 16534945
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 17165322
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.241 min
Delta R.T.: 0.000 min
Response: 20460052
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.217 min
Delta R.T.: 0.000 min
Response: 32471722
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.501 min
Delta R.T.: 0.000 min
Response: 22486979
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.406 min
Delta R.T.: 0.000 min
Response: 37254776
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.787 min
Delta R.T.: 0.000 min
Response: 27163979
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.560 min
Delta R.T.: 0.000 min
Response: 35318401
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.095 min
Delta R.T.: 0.000 min
Response: 19711076
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.036 min
Delta R.T.: 0.000 min
Response: 25439075
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.423 min
Delta R.T.: 0.000 min
Response: 31932860
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.278 min
Delta R.T.: 0.000 min
Response: 57556124
Conc: 500.00 ng/ml