

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061623\
 Data File : P0095770.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2023 05:48
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 17 06:08:05 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Jun 17 06:04:42 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.428	3.633	63021051	35020173	25.208	24.888
2)	SA Decachlor...	10.308	8.674	28564483	48564605	24.556	26.298
Target Compounds							
41)	L9 AR-1268-1	8.745	7.562	23657825	42654380	244.033	260.278
42)	L9 AR-1268-2	8.841	7.627	20201408	37488843	239.779	258.199
43)	L9 AR-1268-3	9.079	7.836	19973095	34009640	249.458	257.634
44)	L9 AR-1268-4	9.518	8.124	7016632	13733864	248.586	261.123
45)	L9 AR-1268-5	9.952	8.417	61772430	95360794	250.906	251.674



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061623\
Data File : PO095770.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2023 05:48
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 17 06:08:05 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061623.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Jun 17 06:04:42 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: 0.000 min
Response: 63021051
Conc: 25.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.633 min
Delta R.T.: 0.000 min
Response: 35020173
Conc: 24.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.308 min
Delta R.T.: 0.002 min
Response: 28564483
Conc: 24.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.674 min
Delta R.T.: 0.000 min
Response: 48564605
Conc: 26.30 ng/ml



#41 AR-1268-1


R.T.: 8.745 min
Delta R.T.: 0.001 min
Response: 23657825
Conc: 244.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.562 min
Delta R.T.: 0.000 min
Response: 42654380
Conc: 260.28 ng/ml




#42 AR-1268-2

R.T.: 8.841 min
Delta R.T.: 0.000 min
Response: 20201408
Conc: 239.78 ng/ml



#42 AR-1268-2


R.T.: 7.627 min
Delta R.T.: 0.000 min
Response: 37488843
Conc: 258.20 ng/ml



#43 AR-1268-3


R.T.: 9.079 min
Delta R.T.: 0.000 min
Response: 19973095
Conc: 249.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.836 min
Delta R.T.: 0.000 min
Response: 34009640
Conc: 257.63 ng/ml




#44 AR-1268-4

R.T.: 9.518 min
Delta R.T.: -0.001 min
Response: 7016632
Conc: 248.59 ng/ml



#44 AR-1268-4


R.T.: 8.124 min
Delta R.T.: 0.000 min
Response: 13733864
Conc: 261.12 ng/ml



#45 AR-1268-5

R.T.: 9.952 min
Delta R.T.: 0.000 min
Response: 61772430
Conc: 250.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.417 min
Delta R.T.: 0.000 min
Response: 95360794
Conc: 251.67 ng/ml