

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061721\
 Data File : P0078892.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 18 Jun 2021 4:42
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 06:55:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Jun 14 05:33:30 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.328	3.605	1843533	709770	21.774	20.503
2)	SA Decachlor...	9.933	8.785	1312889	651383	22.372	20.950
Target Compounds							
9)	L2 AR-1221-2	4.671	0.000	23659	0	35.658	N.D. #
15)	L3 AR-1232-5	5.922	0.000	28507	0	54.552	N.D. #
20)	L4 AR-1242-5	6.601	5.710f	5862	19392	3.898	21.676 #
22)	L5 AR-1248-2	5.922	0.000	28507	0	13.661	N.D. #
24)	L5 AR-1248-4	6.560	0.000	21807	0	8.328	N.D. #
25)	L5 AR-1248-5	6.601	5.710f	5862	19392	2.320	17.591 #
26)	L6 AR-1254-1	6.528	5.710f	70058	19392	25.558	10.976 #
28)	L6 AR-1254-3	7.137	0.000	73811	0	16.953	N.D. #
29)	L6 AR-1254-4	7.411	0.000	5086	0	1.652	N.D. #
39)	L8 AR-1262-4	8.796	0.000	1045	0	0.289	N.D. #
42)	L9 AR-1268-2	8.796	0.000	1045	0	0.145	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061721\
Data File : P0078892.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 18 Jun 2021 4:42
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampled :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 06:55:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061321.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Jun 14 05:33:30 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.328 min
Delta R.T.: -0.003 min
Response: 1843533
Conc: 21.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.605 min
Delta R.T.: -0.002 min
Response: 709770
Conc: 20.50 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.933 min
Delta R.T.: -0.005 min
Response: 1312889
Conc: 22.37 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.785 min
Delta R.T.: -0.005 min
Response: 651383
Conc: 20.95 ng/ml



#9 AR-1221-2


R.T.: 4.671 min
Delta R.T.: -0.006 min
Response: 23659
Conc: 35.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.956 min
Response: 0
Conc: N.D.




#15 AR-1232-5

R.T.: 5.922 min
Delta R.T.: 0.005 min
Response: 28507
Conc: 54.55 ng/ml



#15 AR-1232-5


R.T.: 0.000 min
Exp R.T. : 5.316 min
Response: 0
Conc: N.D.



#20 AR-1242-5


R.T.: 6.601 min
Delta R.T.: 0.012 min
Response: 5862
Conc: 3.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#20 AR-1242-5

R.T.: 5.710 min
Delta R.T.: 0.020 min
Response: 19392
Conc: 21.68 ng/ml




#22 AR-1248-2

R.T.: 5.922 min
Delta R.T.: 0.005 min
Response: 28507
Conc: 13.66 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 5.093 min
Response: 0
Conc: N.D.



#24 AR-1248-4


R.T.: 6.560 min
Delta R.T.: 0.009 min
Response: 21807
Conc: 8.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#24 AR-1248-4

R.T.: 0.000 min
Exp R.T. : 5.316 min
Response: 0
Conc: N.D.




#25 AR-1248-5

R.T.: 6.601 min
Delta R.T.: 0.012 min
Response: 5862
Conc: 2.32 ng/ml



#25 AR-1248-5


R.T.: 5.710 min
Delta R.T.: -0.022 min
Response: 19392
Conc: 17.59 ng/ml



#26 AR-1254-1


R.T.: 6.528 min
Delta R.T.: 0.007 min
Response: 70058
Conc: 25.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#26 AR-1254-1

R.T.: 5.710 min
Delta R.T.: 0.021 min
Response: 19392
Conc: 10.98 ng/ml




#28 AR-1254-3

R.T.: 7.137 min
Delta R.T.: 0.013 min
Response: 73811
Conc: 16.95 ng/ml



#28 AR-1254-3


R.T.: 0.000 min
Exp R.T. : 6.260 min
Response: 0
Conc: N.D.



#29 AR-1254-4


R.T.: 7.411 min
Delta R.T.: -0.006 min
Response: 5086
Conc: 1.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#29 AR-1254-4

R.T.: 0.000 min
Exp R.T. : 6.498 min
Response: 0
Conc: N.D.




#39 AR-1262-4

R.T.: 8.796 min
Delta R.T.: -0.005 min
Response: 1045
Conc: 0.29 ng/ml



#39 AR-1262-4


R.T.: 0.000 min
Exp R.T. : 7.811 min
Response: 0
Conc: N.D.



#42 AR-1268-2

R.T.: 8.796 min
Delta R.T.: -0.007 min
Response: 1045
Conc: 0.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK



#42 AR-1268-2

R.T.: 0.000 min
Exp R.T. : 7.813 min
Response: 0
Conc: N.D.