

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061725\
 Data File : P0111646.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Jun 2025 10:16
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 18 02:01:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061125.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Jun 12 06:25:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	3.677	3.673	310.2E6	279.5E6	53.861	49.758
2)	SA Decachlor...	8.713	8.660	262.5E6	84349606	50.016	47.450
Target Compounds							
16)	L4 AR-1242-1	4.765	4.750	102.5E6	80733831	508.718	496.031
17)	L4 AR-1242-2	4.785	4.768	147.0E6	119.7E6	521.862	506.986
18)	L4 AR-1242-3	4.841	4.944	102.9E6	62830845	517.747	500.288
19)	L4 AR-1242-4	4.961	5.028	81081656	61537871	512.536	502.109
20)	L4 AR-1242-5	5.613	5.548	85373523	75508758	510.557	503.764



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061725\
Data File : PO111646.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Jun 2025 10:16
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 18 02:01:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061125.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Jun 12 06:25:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.677 min
Delta R.T.: 0.000 min
Response: 310206466
Conc: 53.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.673 min
Delta R.T.: 0.000 min
Response: 279504256
Conc: 49.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.713 min
Delta R.T.: 0.000 min
Response: 262545544
Conc: 50.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.660 min
Delta R.T.: -0.002 min
Response: 84349606
Conc: 47.45 ng/ml



#16 AR-1242-1


R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 102478139
Conc: 508.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.750 min
Delta R.T.: 0.000 min
Response: 80733831
Conc: 496.03 ng/ml




#17 AR-1242-2

R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 147043378
Conc: 521.86 ng/ml



#17 AR-1242-2


R.T.: 4.768 min
Delta R.T.: 0.000 min
Response: 119679697
Conc: 506.99 ng/ml



#18 AR-1242-3


R.T.: 4.841 min
Delta R.T.: 0.000 min
Response: 102940256
Conc: 517.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.944 min
Delta R.T.: 0.000 min
Response: 62830845
Conc: 500.29 ng/ml




#19 AR-1242-4

R.T.: 4.961 min
Delta R.T.: 0.000 min
Response: 81081656
Conc: 512.54 ng/ml



#19 AR-1242-4


R.T.: 5.028 min
Delta R.T.: 0.000 min
Response: 61537871
Conc: 502.11 ng/ml



#20 AR-1242-5

R.T.: 5.613 min
Delta R.T.: 0.000 min
Response: 85373523
Conc: 510.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.548 min
Delta R.T.: -0.001 min
Response: 75508758
Conc: 503.76 ng/ml