

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0061924\
 Data File : P0104349.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 19 Jun 2024 08:20
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 19 10:54:02 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0061124.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Jun 12 05:42:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.403	3.546	422.4E6	93535553	50.617	43.384
2)	SA Decachlor...	10.100	8.457	296.9E6	169.7E6	50.687	53.019
Target Compounds							
21)	L5 AR-1248-1	5.563	4.604	89452995	17707477	494.641	391.519
22)	L5 AR-1248-2	5.836	4.835	128.4E6	22039106	500.023	392.244
23)	L5 AR-1248-3	6.039	4.877	137.9E6	25198931	504.730	402.347
24)	L5 AR-1248-4	6.441	5.045	150.8E6	30818869	504.233	402.028
25)	L5 AR-1248-5	6.479	5.430	146.5E6	34935783	507.220	418.442



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO061924\
Data File : PO104349.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 19 Jun 2024 08:20
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 19 10:54:02 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO061124.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Jun 12 05:42:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.403 min
Delta R.T.: 0.002 min
Response: 422426081
Conc: 50.62 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.546 min
Delta R.T.: 0.000 min
Response: 93535553
Conc: 43.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.100 min
Delta R.T.: -0.004 min
Response: 296924841
Conc: 50.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.457 min
Delta R.T.: -0.003 min
Response: 169710742
Conc: 53.02 ng/ml



#21 AR-1248-1


R.T.: 5.563 min
Delta R.T.: 0.001 min
Response: 89452995
Conc: 494.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.604 min
Delta R.T.: -0.001 min
Response: 17707477
Conc: 391.52 ng/ml




#22 AR-1248-2

R.T.: 5.836 min
Delta R.T.: 0.000 min
Response: 128448413
Conc: 500.02 ng/ml



#22 AR-1248-2


R.T.: 4.835 min
Delta R.T.: -0.002 min
Response: 22039106
Conc: 392.24 ng/ml



#23 AR-1248-3


R.T.: 6.039 min
Delta R.T.: 0.000 min
Response: 137886050
Conc: 504.73 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.877 min
Delta R.T.: -0.002 min
Response: 25198931
Conc: 402.35 ng/ml




#24 AR-1248-4

R.T.: 6.441 min
Delta R.T.: 0.000 min
Response: 150803581
Conc: 504.23 ng/ml



#24 AR-1248-4


R.T.: 5.045 min
Delta R.T.: 0.000 min
Response: 30818869
Conc: 402.03 ng/ml



#25 AR-1248-5

R.T.: 6.479 min
Delta R.T.: 0.000 min
Response: 146453286
Conc: 507.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.430 min
Delta R.T.: -0.001 min
Response: 34935783
Conc: 418.44 ng/ml