

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0062022\
 Data File : P0087233.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2022 16:35
 Operator : YP/AJ
 Sample : N3378-01
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 17:59:44 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.493	3.642	44515599	24327362	19.048	21.607
2)	SA Decachlor...	10.348	8.645	20309665	8929356	11.036	11.357
Target Compounds							
3)	L1 AR-1016-1	5.694	4.736	307467	32706	3.673	0.971 #
4)	L1 AR-1016-2	5.701	4.745	265760	46095	2.260	0.991 #
5)	L1 AR-1016-3	5.765	4.927	143137	28645	1.902	1.118 #
6)	L1 AR-1016-4	5.879	4.967	133503	50222	2.228	2.464
7)	L1 AR-1016-5	6.124f	5.186	10310	13435	0.175	0.526 #
8)	L2 AR-1221-1	4.704	3.863	25522745	58207	911.648	4.395 #
9)	L2 AR-1221-2	4.804	3.926	636207	9692661	30.131	976.330 #
10)	L2 AR-1221-3	4.873	4.018	33698	100910	0.522	3.468 #
11)	L3 AR-1232-1	4.873	4.018	33698	100910	0.573	3.782 #
12)	L3 AR-1232-2	5.223	4.745	210.6E6	46095	7275.699	2.126 #
13)	L3 AR-1232-3	5.701	4.927	265760	28645	4.950	2.523 #
14)	L3 AR-1232-4	5.879	5.007	133503	93008	4.971	9.098 #
15)	L3 AR-1232-5	5.957	5.186	89819	13435	4.040	1.192 #
16)	L4 AR-1242-1	5.694	4.736	307467	32706	4.784	1.236 #
17)	L4 AR-1242-2	5.701	4.745	265760	46095	2.908	1.280 #
18)	L4 AR-1242-3	5.765	4.927	143137	28645	2.439	1.433 #
19)	L4 AR-1242-4	5.846	5.007	610941	93008	13.114	4.758 #
20)	L4 AR-1242-5	6.604	5.537	47763	20355	1.007	0.872
21)	L5 AR-1248-1	5.694	4.736	307467	32706	6.337	1.653 #
22)	L5 AR-1248-2	5.957	4.967	89819	50222	1.261	1.848 #
23)	L5 AR-1248-3	6.124f	5.007	10310	93008	0.135	3.255 #
24)	L5 AR-1248-4	6.576	5.186	1070604	13435	12.859	0.408 #
25)	L5 AR-1248-5	6.604	5.578	47763	4358	0.595	0.136 #
26)	L6 AR-1254-1	6.545	5.537	81202	20355	0.931	0.421 #
27)	L6 AR-1254-2	6.757	5.686	138128	14808	1.058	0.347 #
28)	L6 AR-1254-3	7.130	6.089	880136	250831	6.746	3.670 #
29)	L6 AR-1254-4	7.422	6.314	27078	109989	0.283	2.607 #
30)	L6 AR-1254-5	7.838	6.731	336291	34574	3.219	0.569 #
31)	L7 AR-1260-1	7.293	6.212	529541	64572	5.255	1.379 #
32)	L7 AR-1260-2	7.564	6.407	1395793	240952	11.843	4.302 #
33)	L7 AR-1260-3	7.908	6.565	91620	69532	1.081	1.307
34)	L7 AR-1260-4	8.139	7.028	235882	16498	2.316	0.409 #
35)	L7 AR-1260-5	8.474	7.272	84833	51425	0.467	0.548
36)	L8 AR-1262-1	7.908	6.827	91620	50081	0.745	2.070 #
37)	L8 AR-1262-2	8.474	7.272	84833	51425	0.413	0.509
38)	L8 AR-1262-3	8.792	7.551	181718	9689	1.259	0.239 #
39)	L8 AR-1262-4	8.889	7.618	26588	32535	0.377	0.435
40)	L8 AR-1262-5	9.552	8.116	628834	9356	8.137	0.272 #
41)	L9 AR-1268-1	8.792	7.551	181718	9689	0.707	0.081 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062022\
Data File : P0087233.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 16:35
Operator : YP/AJ
Sample : N3378-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 17:59:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42)	L9 AR-1268-2	8.889	7.618	26588	32535	0.112	0.299 #
43)	L9 AR-1268-3	9.128	7.827	77795	20879	0.389	0.227 #
44)	L9 AR-1268-4	9.552	8.116	628834	9356	6.975	0.235 #
45)	L9 AR-1268-5	9.999	8.402	67943	18783	0.106	0.065 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062022\
Data File : P0087233.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 16:35
Operator : YP/AJ
Sample : N3378-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 17:59:44 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.493 min
Delta R.T.: 0.000 min
Response: 44515599
Conc: 19.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.642 min
Delta R.T.: 0.000 min
Response: 24327362
Conc: 21.61 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.348 min
Delta R.T.: -0.008 min
Response: 20309665
Conc: 11.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: -0.013 min
Response: 8929356
Conc: 11.36 ng/ml



#3 AR-1016-1


R.T.: 5.694 min
Delta R.T.: 0.016 min
Response: 307467
Conc: 3.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.736 min
Delta R.T.: 0.004 min
Response: 32706
Conc: 0.97 ng/ml




#4 AR-1016-2

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 265760
Conc: 2.26 ng/ml



#4 AR-1016-2


R.T.: 4.745 min
Delta R.T.: -0.006 min
Response: 46095
Conc: 0.99 ng/ml



#5 AR-1016-3


R.T.: 5.765 min
Delta R.T.: 0.001 min
Response: 143137
Conc: 1.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.927 min
Delta R.T.: 0.000 min
Response: 28645
Conc: 1.12 ng/ml




#6 AR-1016-4

R.T.: 5.879 min
Delta R.T.: 0.016 min
Response: 133503
Conc: 2.23 ng/ml



#6 AR-1016-4


R.T.: 4.967 min
Delta R.T.: -0.003 min
Response: 50222
Conc: 2.46 ng/ml



#7 AR-1016-5


R.T.: 6.124 min
Delta R.T.: -0.037 min
Response: 10310
Conc: 0.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.186 min
Delta R.T.: 0.003 min
Response: 13435
Conc: 0.53 ng/ml




#8 AR-1221-1

R.T.: 4.704 min
Delta R.T.: 0.000 min
Response: 25522745
Conc: 911.65 ng/ml



#8 AR-1221-1


R.T.: 3.863 min
Delta R.T.: 0.005 min
Response: 58207
Conc: 4.40 ng/ml



#9 AR-1221-2


R.T.: 4.804 min
Delta R.T.: 0.013 min
Response: 636207
Conc: 30.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.926 min
Delta R.T.: -0.018 min
Response: 9692661
Conc: 976.33 ng/ml




#10 AR-1221-3

R.T.: 4.873 min
Delta R.T.: 0.004 min
Response: 33698
Conc: 0.52 ng/ml



#10 AR-1221-3


R.T.: 4.018 min
Delta R.T.: -0.002 min
Response: 100910
Conc: 3.47 ng/ml



#11 AR-1232-1


R.T.: 4.873 min
Delta R.T.: 0.004 min
Response: 33698
Conc: 0.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.018 min
Delta R.T.: -0.003 min
Response: 100910
Conc: 3.78 ng/ml




#12 AR-1232-2

R.T.: 5.223 min
Delta R.T.: 0.012 min
Response: 210583829
Conc: 7275.70 ng/ml



#12 AR-1232-2


R.T.: 4.745 min
Delta R.T.: -0.006 min
Response: 46095
Conc: 2.13 ng/ml



#13 AR-1232-3


R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 265760
Conc: 4.95 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.927 min
Delta R.T.: 0.000 min
Response: 28645
Conc: 2.52 ng/ml




#14 AR-1232-4

R.T.: 5.879 min
Delta R.T.: 0.015 min
Response: 133503
Conc: 4.97 ng/ml



#14 AR-1232-4


R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 93008
Conc: 9.10 ng/ml



#15 AR-1232-5


R.T.: 5.957 min
Delta R.T.: 0.002 min
Response: 89819
Conc: 4.04 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.186 min
Delta R.T.: 0.002 min
Response: 13435
Conc: 1.19 ng/ml




#16 AR-1242-1

R.T.: 5.694 min
Delta R.T.: 0.017 min
Response: 307467
Conc: 4.78 ng/ml



#16 AR-1242-1


R.T.: 4.736 min
Delta R.T.: 0.004 min
Response: 32706
Conc: 1.24 ng/ml



#17 AR-1242-2


R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 265760
Conc: 2.91 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.745 min
Delta R.T.: -0.006 min
Response: 46095
Conc: 1.28 ng/ml




#18 AR-1242-3

R.T.: 5.765 min
Delta R.T.: 0.003 min
Response: 143137
Conc: 2.44 ng/ml



#18 AR-1242-3


R.T.: 4.927 min
Delta R.T.: 0.000 min
Response: 28645
Conc: 1.43 ng/ml



#19 AR-1242-4


R.T.: 5.846 min
Delta R.T.: -0.016 min
Response: 610941
Conc: 13.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.007 min
Delta R.T.: -0.005 min
Response: 93008
Conc: 4.76 ng/ml




#20 AR-1242-5

R.T.: 6.604 min
Delta R.T.: -0.001 min
Response: 47763
Conc: 1.01 ng/ml



#20 AR-1242-5


R.T.: 5.537 min
Delta R.T.: 0.003 min
Response: 20355
Conc: 0.87 ng/ml



#21 AR-1248-1


R.T.: 5.694 min
Delta R.T.: 0.017 min
Response: 307467
Conc: 6.34 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.736 min
Delta R.T.: 0.004 min
Response: 32706
Conc: 1.65 ng/ml




#22 AR-1248-2

R.T.: 5.957 min
Delta R.T.: 0.002 min
Response: 89819
Conc: 1.26 ng/ml



#22 AR-1248-2


R.T.: 4.967 min
Delta R.T.: -0.003 min
Response: 50222
Conc: 1.85 ng/ml



#23 AR-1248-3


R.T.: 6.124 min
Delta R.T.: -0.036 min
Response: 10310
Conc: 0.14 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.007 min
Delta R.T.: -0.004 min
Response: 93008
Conc: 3.25 ng/ml




#24 AR-1248-4

R.T.: 6.576 min
Delta R.T.: 0.010 min
Response: 1070604
Conc: 12.86 ng/ml



#24 AR-1248-4


R.T.: 5.186 min
Delta R.T.: 0.003 min
Response: 13435
Conc: 0.41 ng/ml



#25 AR-1248-5


R.T.: 6.604 min
Delta R.T.: -0.003 min
Response: 47763
Conc: 0.60 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.578 min
Delta R.T.: 0.003 min
Response: 4358
Conc: 0.14 ng/ml




#26 AR-1254-1

R.T.: 6.545 min
Delta R.T.: 0.005 min
Response: 81202
Conc: 0.93 ng/ml



#26 AR-1254-1


R.T.: 5.537 min
Delta R.T.: 0.002 min
Response: 20355
Conc: 0.42 ng/ml



#27 AR-1254-2


R.T.: 6.757 min
Delta R.T.: -0.004 min
Response: 138128
Conc: 1.06 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.686 min
Delta R.T.: 0.004 min
Response: 14808
Conc: 0.35 ng/ml




#28 AR-1254-3

R.T.: 7.130 min
Delta R.T.: 0.000 min
Response: 880136
Conc: 6.75 ng/ml



#28 AR-1254-3


R.T.: 6.089 min
Delta R.T.: 0.004 min
Response: 250831
Conc: 3.67 ng/ml



#29 AR-1254-4


R.T.: 7.422 min
Delta R.T.: 0.006 min
Response: 27078
Conc: 0.28 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.314 min
Delta R.T.: 0.000 min
Response: 109989
Conc: 2.61 ng/ml




#30 AR-1254-5

R.T.: 7.838 min
Delta R.T.: -0.002 min
Response: 336291
Conc: 3.22 ng/ml



#30 AR-1254-5


R.T.: 6.731 min
Delta R.T.: 0.000 min
Response: 34574
Conc: 0.57 ng/ml



#31 AR-1260-1


R.T.: 7.293 min
Delta R.T.: -0.004 min
Response: 529541
Conc: 5.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.212 min
Delta R.T.: -0.005 min
Response: 64572
Conc: 1.38 ng/ml




#32 AR-1260-2

R.T.: 7.564 min
Delta R.T.: 0.009 min
Response: 1395793
Conc: 11.84 ng/ml



#32 AR-1260-2


R.T.: 6.407 min
Delta R.T.: 0.003 min
Response: 240952
Conc: 4.30 ng/ml



#33 AR-1260-3


R.T.: 7.908 min
Delta R.T.: -0.009 min
Response: 91620
Conc: 1.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.565 min
Delta R.T.: 0.007 min
Response: 69532
Conc: 1.31 ng/ml




#34 AR-1260-4

R.T.: 8.139 min
Delta R.T.: -0.007 min
Response: 235882
Conc: 2.32 ng/ml



#34 AR-1260-4


R.T.: 7.028 min
Delta R.T.: -0.002 min
Response: 16498
Conc: 0.41 ng/ml



#35 AR-1260-5


R.T.: 8.474 min
Delta R.T.: 0.000 min
Response: 84833
Conc: 0.47 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.272 min
Delta R.T.: 0.000 min
Response: 51425
Conc: 0.55 ng/ml




#36 AR-1262-1

R.T.: 7.908 min
Delta R.T.: -0.006 min
Response: 91620
Conc: 0.75 ng/ml



#36 AR-1262-1


R.T.: 6.827 min
Delta R.T.: 0.005 min
Response: 50081
Conc: 2.07 ng/ml



#37 AR-1262-2


R.T.: 8.474 min
Delta R.T.: 0.002 min
Response: 84833
Conc: 0.41 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.272 min
Delta R.T.: 0.004 min
Response: 51425
Conc: 0.51 ng/ml




#38 AR-1262-3

R.T.: 8.792 min
Delta R.T.: -0.003 min
Response: 181718
Conc: 1.26 ng/ml



#38 AR-1262-3


R.T.: 7.551 min
Delta R.T.: -0.001 min
Response: 9689
Conc: 0.24 ng/ml



#39 AR-1262-4


R.T.: 8.889 min
Delta R.T.: 0.001 min
Response: 26588
Conc: 0.38 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.618 min
Delta R.T.: 0.003 min
Response: 32535
Conc: 0.43 ng/ml




#40 AR-1262-5

R.T.: 9.552 min
Delta R.T.: -0.016 min
Response: 628834
Conc: 8.14 ng/ml



#40 AR-1262-5


R.T.: 8.116 min
Delta R.T.: 0.006 min
Response: 9356
Conc: 0.27 ng/ml



#41 AR-1268-1


R.T.: 8.792 min
Delta R.T.: 0.000 min
Response: 181718
Conc: 0.71 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.551 min
Delta R.T.: -0.002 min
Response: 9689
Conc: 0.08 ng/ml




#42 AR-1268-2

R.T.: 8.889 min
Delta R.T.: -0.001 min
Response: 26588
Conc: 0.11 ng/ml



#42 AR-1268-2


R.T.: 7.618 min
Delta R.T.: 0.002 min
Response: 32535
Conc: 0.30 ng/ml



#43 AR-1268-3


R.T.: 9.128 min
Delta R.T.: 0.002 min
Response: 77795
Conc: 0.39 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.827 min
Delta R.T.: 0.003 min
Response: 20879
Conc: 0.23 ng/ml




#44 AR-1268-4

R.T.: 9.552 min
Delta R.T.: -0.013 min
Response: 628834
Conc: 6.98 ng/ml



#44 AR-1268-4


R.T.: 8.116 min
Delta R.T.: 0.007 min
Response: 9356
Conc: 0.24 ng/ml



#45 AR-1268-5

R.T.: 9.999 min
Delta R.T.: 0.000 min
Response: 67943
Conc: 0.11 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.402 min
Delta R.T.: 0.002 min
Response: 18783
Conc: 0.07 ng/ml