

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0062022\
 Data File : P0087236.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Jun 2022 17:25
 Operator : YP/AJ
 Sample : N3398-12
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jun 20 18:01:13 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0060922.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Jun 10 04:33:49 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.494	3.643	31442978	16677501	13.454	14.812
2)	SA Decachlor...	10.349	8.645	19398563	8192660	10.541	10.420
Target Compounds							
3)	L1 AR-1016-1	5.690	4.731	244159	29596	2.917	0.879 #
4)	L1 AR-1016-2	5.705	4.747	1571344	247485	13.363	5.320 #
5)	L1 AR-1016-3	5.763	4.925	1463941	50299	19.448	1.963 #
6)	L1 AR-1016-4	5.860	4.968	719175	13827	12.003	0.678 #
7)	L1 AR-1016-5	6.164	5.184	58328	3617	0.992	0.142 #
8)	L2 AR-1221-1	4.703	3.856	1550739	20903	55.391	1.578 #
9)	L2 AR-1221-2	4.782	3.956	44008	76158	2.084	7.671 #
10)	L2 AR-1221-3	4.872	4.017	538692	252137	8.340	8.665
11)	L3 AR-1232-1	4.872	4.017	538692	252137	9.161	9.451
12)	L3 AR-1232-2	5.218	4.747	5379304	247485	185.856	11.412 #
13)	L3 AR-1232-3	5.705	4.925	1571344	50299	29.265	4.430 #
14)	L3 AR-1232-4	5.860	5.011	719175	47986	26.778	4.694 #
15)	L3 AR-1232-5	5.944	5.184	152801	3617	6.874	0.321 #
16)	L4 AR-1242-1	5.690	4.731	244159	29596	3.799	1.118 #
17)	L4 AR-1242-2	5.705	4.747	1571344	247485	17.197	6.871 #
18)	L4 AR-1242-3	5.763	4.925	1463941	50299	24.941	2.517 #
19)	L4 AR-1242-4	5.860	5.011	719175	47986	15.437	2.455 #
20)	L4 AR-1242-5	6.607	5.534	35300	55817	0.744	2.391 #
21)	L5 AR-1248-1	5.690	4.731	244159	29596	5.032	1.496 #
22)	L5 AR-1248-2	5.944	4.968	152801	13827	2.145	0.509 #
23)	L5 AR-1248-3	6.164	5.011	58328	47986	0.764	1.679 #
24)	L5 AR-1248-4	6.563	5.184	35315	3617	0.424	0.110 #
25)	L5 AR-1248-5	6.607	5.576	35300	83399	0.440	2.609 #
26)	L6 AR-1254-1	6.538	5.534	75398	55817	0.865	1.153 #
27)	L6 AR-1254-2	6.755	5.682	48703	74262	0.373	1.741 #
28)	L6 AR-1254-3	7.126	6.080	342885	100495	2.628	1.470 #
29)	L6 AR-1254-4	7.414	6.314	37212	81233	0.389	1.925 #
30)	L6 AR-1254-5	7.842	6.728	27703	54336	0.265	0.895 #
31)	L7 AR-1260-1	7.296	6.221	67660	111079	0.671	2.371 #
32)	L7 AR-1260-2	7.553	6.400	35212	344835	0.299	6.156 #
33)	L7 AR-1260-3	7.922	6.560	105222	57816	1.241	1.087
34)	L7 AR-1260-4	8.146	7.032	201820	26096	1.982	0.648 #
35)	L7 AR-1260-5	8.471	7.274	127428	27240	0.701	0.290 #
36)	L8 AR-1262-1	7.922	6.822	105222	24151	0.856	0.998
37)	L8 AR-1262-2	8.471	7.263	127428	34164	0.621	0.338 #
38)	L8 AR-1262-3	8.795	7.552	57926	145795	0.401	3.600 #
39)	L8 AR-1262-4	8.890	7.617	52092	35755	0.738	0.478 #
40)	L8 AR-1262-5	9.561	8.108	77387	16336	1.001	0.475 #
41)	L9 AR-1268-1	8.795	7.552	57926	145795	0.225	1.221 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062022\
Data File : P0087236.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 17:25
Operator : YP/AJ
Sample : N3398-12
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 18:01:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.890	7.617	52092	35755	0.220	0.328 #
43) L9	AR-1268-3	9.122	7.824	59639	8303	0.298	0.090 #
44) L9	AR-1268-4	9.561	8.108	77387	16336	0.858	0.411 #
45) L9	AR-1268-5	9.996	8.400	43118	10419	0.067	0.036 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0062022\
Data File : P0087236.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Jun 2022 17:25
Operator : YP/AJ
Sample : N3398-12
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jun 20 18:01:13 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0060922.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Jun 10 04:33:49 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.494 min
Delta R.T.: 0.000 min
Response: 31442978
Conc: 13.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.643 min
Delta R.T.: 0.002 min
Response: 16677501
Conc: 14.81 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.349 min
Delta R.T.: -0.006 min
Response: 19398563
Conc: 10.54 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.645 min
Delta R.T.: -0.012 min
Response: 8192660
Conc: 10.42 ng/ml



#3 AR-1016-1


R.T.: 5.690 min
Delta R.T.: 0.012 min
Response: 244159
Conc: 2.92 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 4.731 min
Delta R.T.: -0.001 min
Response: 29596
Conc: 0.88 ng/ml




#4 AR-1016-2

R.T.: 5.705 min
Delta R.T.: 0.005 min
Response: 1571344
Conc: 13.36 ng/ml



#4 AR-1016-2


R.T.: 4.747 min
Delta R.T.: -0.004 min
Response: 247485
Conc: 5.32 ng/ml



#5 AR-1016-3


R.T.: 5.763 min
Delta R.T.: 0.000 min
Response: 1463941
Conc: 19.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 4.925 min
Delta R.T.: -0.002 min
Response: 50299
Conc: 1.96 ng/ml




#6 AR-1016-4

R.T.: 5.860 min
Delta R.T.: -0.003 min
Response: 719175
Conc: 12.00 ng/ml



#6 AR-1016-4


R.T.: 4.968 min
Delta R.T.: -0.002 min
Response: 13827
Conc: 0.68 ng/ml



#7 AR-1016-5


R.T.: 6.164 min
Delta R.T.: 0.003 min
Response: 58328
Conc: 0.99 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.184 min
Delta R.T.: 0.000 min
Response: 3617
Conc: 0.14 ng/ml




#8 AR-1221-1

R.T.: 4.703 min
Delta R.T.: 0.000 min
Response: 1550739
Conc: 55.39 ng/ml



#8 AR-1221-1


R.T.: 3.856 min
Delta R.T.: -0.003 min
Response: 20903
Conc: 1.58 ng/ml



#9 AR-1221-2


R.T.: 4.782 min
Delta R.T.: -0.010 min
Response: 44008
Conc: 2.08 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 3.956 min
Delta R.T.: 0.012 min
Response: 76158
Conc: 7.67 ng/ml




#10 AR-1221-3

R.T.: 4.872 min
Delta R.T.: 0.003 min
Response: 538692
Conc: 8.34 ng/ml



#10 AR-1221-3


R.T.: 4.017 min
Delta R.T.: -0.003 min
Response: 252137
Conc: 8.67 ng/ml



#11 AR-1232-1


R.T.: 4.872 min
Delta R.T.: 0.003 min
Response: 538692
Conc: 9.16 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 4.017 min
Delta R.T.: -0.004 min
Response: 252137
Conc: 9.45 ng/ml




#12 AR-1232-2

R.T.: 5.218 min
Delta R.T.: 0.007 min
Response: 5379304
Conc: 185.86 ng/ml



#12 AR-1232-2


R.T.: 4.747 min
Delta R.T.: -0.005 min
Response: 247485
Conc: 11.41 ng/ml



#13 AR-1232-3


R.T.: 5.705 min
Delta R.T.: 0.005 min
Response: 1571344
Conc: 29.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 4.925 min
Delta R.T.: -0.002 min
Response: 50299
Conc: 4.43 ng/ml




#14 AR-1232-4

R.T.: 5.860 min
Delta R.T.: -0.004 min
Response: 719175
Conc: 26.78 ng/ml



#14 AR-1232-4


R.T.: 5.011 min
Delta R.T.: 0.000 min
Response: 47986
Conc: 4.69 ng/ml



#15 AR-1232-5


R.T.: 5.944 min
Delta R.T.: -0.011 min
Response: 152801
Conc: 6.87 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 5.184 min
Delta R.T.: 0.000 min
Response: 3617
Conc: 0.32 ng/ml




#16 AR-1242-1

R.T.: 5.690 min
Delta R.T.: 0.013 min
Response: 244159
Conc: 3.80 ng/ml



#16 AR-1242-1


R.T.: 4.731 min
Delta R.T.: -0.001 min
Response: 29596
Conc: 1.12 ng/ml



#17 AR-1242-2


R.T.: 5.705 min
Delta R.T.: 0.005 min
Response: 1571344
Conc: 17.20 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 4.747 min
Delta R.T.: -0.004 min
Response: 247485
Conc: 6.87 ng/ml




#18 AR-1242-3

R.T.: 5.763 min
Delta R.T.: 0.001 min
Response: 1463941
Conc: 24.94 ng/ml



#18 AR-1242-3


R.T.: 4.925 min
Delta R.T.: -0.002 min
Response: 50299
Conc: 2.52 ng/ml



#19 AR-1242-4


R.T.: 5.860 min
Delta R.T.: -0.002 min
Response: 719175
Conc: 15.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 5.011 min
Delta R.T.: 0.000 min
Response: 47986
Conc: 2.46 ng/ml




#20 AR-1242-5

R.T.: 6.607 min
Delta R.T.: 0.002 min
Response: 35300
Conc: 0.74 ng/ml



#20 AR-1242-5


R.T.: 5.534 min
Delta R.T.: 0.000 min
Response: 55817
Conc: 2.39 ng/ml



#21 AR-1248-1


R.T.: 5.690 min
Delta R.T.: 0.013 min
Response: 244159
Conc: 5.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 4.731 min
Delta R.T.: -0.001 min
Response: 29596
Conc: 1.50 ng/ml




#22 AR-1248-2

R.T.: 5.944 min
Delta R.T.: -0.011 min
Response: 152801
Conc: 2.14 ng/ml



#22 AR-1248-2


R.T.: 4.968 min
Delta R.T.: -0.002 min
Response: 13827
Conc: 0.51 ng/ml



#23 AR-1248-3


R.T.: 6.164 min
Delta R.T.: 0.003 min
Response: 58328
Conc: 0.76 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 5.011 min
Delta R.T.: 0.000 min
Response: 47986
Conc: 1.68 ng/ml




#24 AR-1248-4

R.T.: 6.563 min
Delta R.T.: -0.003 min
Response: 35315
Conc: 0.42 ng/ml



#24 AR-1248-4


R.T.: 5.184 min
Delta R.T.: 0.001 min
Response: 3617
Conc: 0.11 ng/ml



#25 AR-1248-5


R.T.: 6.607 min
Delta R.T.: 0.000 min
Response: 35300
Conc: 0.44 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.576 min
Delta R.T.: 0.001 min
Response: 83399
Conc: 2.61 ng/ml




#26 AR-1254-1

R.T.: 6.538 min
Delta R.T.: -0.002 min
Response: 75398
Conc: 0.86 ng/ml



#26 AR-1254-1


R.T.: 5.534 min
Delta R.T.: -0.001 min
Response: 55817
Conc: 1.15 ng/ml



#27 AR-1254-2


R.T.: 6.755 min
Delta R.T.: -0.006 min
Response: 48703
Conc: 0.37 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.682 min
Delta R.T.: 0.000 min
Response: 74262
Conc: 1.74 ng/ml




#28 AR-1254-3

R.T.: 7.126 min
Delta R.T.: -0.004 min
Response: 342885
Conc: 2.63 ng/ml



#28 AR-1254-3


R.T.: 6.080 min
Delta R.T.: -0.005 min
Response: 100495
Conc: 1.47 ng/ml



#29 AR-1254-4


R.T.: 7.414 min
Delta R.T.: -0.002 min
Response: 37212
Conc: 0.39 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.314 min
Delta R.T.: 0.001 min
Response: 81233
Conc: 1.93 ng/ml




#30 AR-1254-5

R.T.: 7.842 min
Delta R.T.: 0.003 min
Response: 27703
Conc: 0.27 ng/ml



#30 AR-1254-5


R.T.: 6.728 min
Delta R.T.: -0.003 min
Response: 54336
Conc: 0.89 ng/ml



#31 AR-1260-1


R.T.: 7.296 min
Delta R.T.: 0.000 min
Response: 67660
Conc: 0.67 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.221 min
Delta R.T.: 0.004 min
Response: 111079
Conc: 2.37 ng/ml




#32 AR-1260-2

R.T.: 7.553 min
Delta R.T.: -0.002 min
Response: 35212
Conc: 0.30 ng/ml



#32 AR-1260-2


R.T.: 6.400 min
Delta R.T.: -0.005 min
Response: 344835
Conc: 6.16 ng/ml



#33 AR-1260-3


R.T.: 7.922 min
Delta R.T.: 0.005 min
Response: 105222
Conc: 1.24 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.560 min
Delta R.T.: 0.001 min
Response: 57816
Conc: 1.09 ng/ml




#34 AR-1260-4

R.T.: 8.146 min
Delta R.T.: 0.000 min
Response: 201820
Conc: 1.98 ng/ml



#34 AR-1260-4


R.T.: 7.032 min
Delta R.T.: 0.002 min
Response: 26096
Conc: 0.65 ng/ml



#35 AR-1260-5


R.T.: 8.471 min
Delta R.T.: -0.004 min
Response: 127428
Conc: 0.70 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.274 min
Delta R.T.: 0.003 min
Response: 27240
Conc: 0.29 ng/ml




#36 AR-1262-1

R.T.: 7.922 min
Delta R.T.: 0.007 min
Response: 105222
Conc: 0.86 ng/ml



#36 AR-1262-1


R.T.: 6.822 min
Delta R.T.: 0.000 min
Response: 24151
Conc: 1.00 ng/ml



#37 AR-1262-2


R.T.: 8.471 min
Delta R.T.: 0.000 min
Response: 127428
Conc: 0.62 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 7.263 min
Delta R.T.: -0.005 min
Response: 34164
Conc: 0.34 ng/ml




#38 AR-1262-3

R.T.: 8.795 min
Delta R.T.: 0.000 min
Response: 57926
Conc: 0.40 ng/ml



#38 AR-1262-3


R.T.: 7.552 min
Delta R.T.: 0.000 min
Response: 145795
Conc: 3.60 ng/ml



#39 AR-1262-4


R.T.: 8.890 min
Delta R.T.: 0.003 min
Response: 52092
Conc: 0.74 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.617 min
Delta R.T.: 0.001 min
Response: 35755
Conc: 0.48 ng/ml




#40 AR-1262-5

R.T.: 9.561 min
Delta R.T.: -0.006 min
Response: 77387
Conc: 1.00 ng/ml



#40 AR-1262-5


R.T.: 8.108 min
Delta R.T.: -0.002 min
Response: 16336
Conc: 0.47 ng/ml



#41 AR-1268-1


R.T.: 8.795 min
Delta R.T.: 0.003 min
Response: 57926
Conc: 0.23 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.552 min
Delta R.T.: -0.001 min
Response: 145795
Conc: 1.22 ng/ml




#42 AR-1268-2

R.T.: 8.890 min
Delta R.T.: 0.000 min
Response: 52092
Conc: 0.22 ng/ml



#42 AR-1268-2


R.T.: 7.617 min
Delta R.T.: 0.000 min
Response: 35755
Conc: 0.33 ng/ml



#43 AR-1268-3


R.T.: 9.122 min
Delta R.T.: -0.004 min
Response: 59639
Conc: 0.30 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 7.824 min
Delta R.T.: 0.000 min
Response: 8303
Conc: 0.09 ng/ml




#44 AR-1268-4

R.T.: 9.561 min
Delta R.T.: -0.003 min
Response: 77387
Conc: 0.86 ng/ml



#44 AR-1268-4


R.T.: 8.108 min
Delta R.T.: -0.002 min
Response: 16336
Conc: 0.41 ng/ml



#45 AR-1268-5

R.T.: 9.996 min
Delta R.T.: -0.002 min
Response: 43118
Conc: 0.07 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.400 min
Delta R.T.: 0.000 min
Response: 10419
Conc: 0.04 ng/ml